



723-3 2025-04-07



For research use only

HCV REAL-TIME PCR Kit

INSTRUCTION FOR USE



"DNA-Technology Research & Production", LLC,

142281, Russia,

Moscow Region, Protvino,

Zheleznodorozhnaya Street, 20

Phone/fax: +7(495) 640.17.71

E-mail: info@dna-technology.com

<https://www.dna-technology.com>

Customer service department

E-mail: hotline@dna-technology.ru

REF

R3-P613-S3/9ER
R3-P613-23/9ER

VER

723-3.2025.04.07

TABLE OF CONTENTS

1. INTENDED USE	3
2. METHOD	3
3. CONTENT	4
4. REAGENTS AND EQUIPMENT REQUIRED BUT NOT PROVIDED	4
5. TRANSPORT AND STORAGE CONDITIONS	6
6. WARNINGS AND PRECAUTIONS	6
7. SAMPLES	8
8. PROCEDURE	8
9. CONTROLS	11
10. DATA ANALYSIS	12
11. SPECIFICATIONS	13
12. TROUBLESHOOTING	14
13. QUALITY CONTROL	15
14. KEY TO SYMBOLS	16

1. INTENDED USE

The **HCV REAL-TIME PCR Kit** is an in vitro Nucleic Acid Test (NAT) – pathogen-detection-based product. The **HCV REAL-TIME PCR Kit** is designed for qualitative detection of hepatitis C virus RNA in human biological material (blood plasma, blood serum) by real-time RT-PCR.

Indications for the test: symptoms of infection, HCV screening.

The application of the kit does not depend on population and demographic aspects. There are no contradictions for use of the **HCV REAL-TIME PCR Kit**.

The **HCV REAL-TIME PCR Kit** can be used in research practice.

Potential users: qualified personnel trained in molecular research methods and rules of work in the laboratory.

It is necessary to apply the kit only as directed in this instruction for use.

2. METHOD

Method: reverse transcription of RNA followed by amplification of synthesized cDNA fragments by real-time polymerase chain reaction (RT-PCR); qualitative analysis.

Method principle: based on the use of reverse transcription of RNA and subsequent amplification of cDNA, which consists of repeated cycles of temperature denaturation of DNA, annealing of primers with complementary sequences and subsequent elongation of polynucleotide chains by Taq-polymerase.

The RNA reverse transcription stage and PCR amplification of cDNA stage are performed in one test tube.

To increase the sensitivity and specificity of the amplification reaction, the use of a hot-start is provided. Hot-start is provided by the PCR-mix preparation consisting of two layers separated by a layer of paraffin. The polymerase chain reaction starts only when paraffin is melted. It excludes non-specific anchoring of primers to DNA target at lower temperatures. Additionally, when the reaction is finished and the tubes cool down, paraffin provides sealing of the mix and additional protection from contamination with amplification products.

The **HCV REAL-TIME PCR Kit** includes the internal control (RNA-IC “A”), which is added into the test samples at RNA extraction stage and is intended to assess the quality of the RNA extraction, reverse transcription and polymerase chain reaction.

DNA probes, each containing a fluorescent label and a fluorescence quencher, are introduced into the PCR-mix. When a specific product is formed, the DNA probe is destroyed and the quenching agent stops affecting the fluorescent label, which leads to an increase in the fluorescence level. The number of destroyed probes (hence the fluorescence level) increases in proportion to the number of specific amplicons formed, and the fluorescence level is measured at each amplification cycle.

The DNA probes used to detect the amplification product of the desired HCV cDNA include the Fam fluorescent label. The DNA probes used to detect the product of the internal control include the Hex fluorescent label. Table 1 shows the detection channels of amplification products.

Table 1. Detection channels of amplification products

Fam	Hex	Rox	Cy5	Cy5.5
RNA HCV	IC*	-	-	-

* Internal control RNA-IC “A”

The automatic analysis is available on “DNA-Technology” made instruments: DTLite or DTprime REAL-TIME Thermal Cyclers for **HCV REAL-TIME PCR Kit** (see the catalogue at <https://www.dna-technology.com> to see available supply options). The current version of the software is available for download at <https://www.dna-technology.com/software>.

The **HCV REAL-TIME PCR Kit** is also approved for use with CFX96 (Bio-Rad) and Applied Biosystems QuantStudio 5 real-time thermal cyclers.

3. CONTENT

The **HCV REAL-TIME PCR Kit** content is represented in Table 2.

Table 2. The **HCV REAL-TIME PCR Kit** content, package S (standard) for R3-P613-S3/9ER and R3-P613-23/9ER

Reagent	Description	Total volume	Amount
Paraffin sealed PCR-mix	Colorless or pink transparent liquid under waxy white fraction	1.44 mL (15 µL in each)	12 strips of 8 or 96 individual tubes
Enzyme Taq/RT	Colorless transparent viscous liquid	55 µL	1 tube
RT-PCR-buffer "V"	Colorless transparent liquid	1.62 mL (810 µL in each)	2 tubes
Internal control RNA-IC "A"	Colorless transparent liquid	2.0 mL (1.0 mL in each)	2 tubes
Positive control*	Colorless transparent liquid	130 µL	1 tube
Strip caps**	12 strips of 8		

* - marking as C+ is allowed

** - for package S, strips (**REF** R3-P613-S3/9ER)

All components are ready to use and do not require additional preparation for operation.

The **HCV REAL-TIME PCR Kit** is intended for single use and designed for 96 tests (no more than 6 runs), including analysis of test samples, negative controls and positive controls.

4. REAGENTS AND EQUIPMENT REQUIRED BUT NOT PROVIDED

The following equipment, reagents and consumables are required for the procedure:

- UV PCR cabinet;
- Real-time detecting thermal cycler¹;
- Vortex mixer;
- Vortex rotor for 0.2 mL stripped tubes (only for package S, strips);
- Refrigerator with freezer;
- PCR tube rack for 0.2 mL tubes or strips;
- Tube rack for 1.5 mL tubes;
- Tube rack for 2.0 mL tubes;
- Single channel pipettes (dispensers covering 2.0-1,000 µL volume range);
- RNase and DNase free filtered pipette tips (volume 20 µL, 200 µL, 1,000 µL);
- Pipette stand;
- RNase and DNase free 1.5 mL microcentrifuge tubes with caps;
- RNase and DNase free 2.0 mL microcentrifuge tubes with caps;
- Powder-free surgical gloves;
- Container for used pipette tips, tubes and other consumables;

¹ Hereinafter – detecting thermal cycler; the required parameters for detecting thermal cyclers are listed below.

- Nucleic acid extraction kit (**PREP-NA-ULTRA** ^{REF} P-017-N/1INT; **PREP-MB-ULTRA** ^{REF} P-132-A/9ER, P-130-N/9ER, P-133-P/9ER, P-134-P/9ER; **PREP-MB-LITE** ^{REF} P-136-A/9ER, P-136-N/9ER, P-136-P/9ER, P-137-P/9ER are recommended).

The following detecting thermal cyclers are recommended:

- **DTprime** (**DTprime *M*** modification), “DNA-Technology R&P”, LLC, Russia;
- **DTlite** (**DTlite *S*** modification), “DNA-Technology R&P”, LLC, Russia;
- **CFX96** (Optical Reaction Module CFX96), Bio-Rad Laboratories, Inc.; USA;
- **Applied Biosystems QuantStudio 5**, Life Technologies Holdings Pte. Ltd., Singapore.

For the use of detecting thermal cyclers other than those listed above, please consult the reagent kit manufacturer for approval.

Software:

The most recent version of the DT thermal cyclers software can be downloaded from <https://www.dna-technology.com/software>.

The OS supported: all versions of Windows starting from 7.

5. TRANSPORT AND STORAGE CONDITIONS

Expiry date – 12 months from the date of production.

All components of **HCV REAL-TIME PCR Kit**, except Enzyme Taq/RT, must be stored at temperatures from 2 °C to 8 °C over the storage period. The PCR-mix for amplification must be stored at temperatures from 2 °C to 8 °C and out of light over the storage period.

The Enzyme Taq/RT must be stored at temperatures from minus 22 °C to minus 18 °C over the storage period.

Transport of the kit, except the Enzyme Taq/RT, is allowed in thermobox with ice packs by all types of roofed transport at temperatures from 2 °C to 25 °C but no more than 5 days and should be stored at temperatures from 2 °C to 8 °C immediately on receipt.

The **HCV REAL-TIME PCR Kit** must be transported in thermoboxes with ice packs by all types of roofed transport at temperatures inside the thermoboxes corresponding to storage conditions of the kit components.

It is allowed to transport the kit, except for Enzyme Taq/RT, in thermoboxes with ice packs by all types of roofed transport at the temperature inside the thermoboxes from 2 °C to 25 °C for no longer than 5 days.

It is allowed to transport the Enzyme Taq/RT in thermoboxes with ice packs by all types of roofed transport at the temperature inside the thermoboxes up to 25 °C for no longer than 5 days.

Shelf-life of the kit following the first opening of the primary container:

- All components of the kit, except for Enzyme Taq/RT, must be stored in a refrigerator or a cooling chamber at temperatures from 2 °C to 8 °C over the storage period;
- PCR-mix must be stored in a refrigerator or a cooling chamber at temperatures from 2 °C to 8 °C and out of light over the storage period;
- Enzyme Taq/RT must be stored in a freezer at temperatures from minus 22 °C to minus 18 °C over the storage period.

The kits stored under undue conditions should not be used.

An expired **HCV REAL-TIME PCR Kit** should not be used.

We strongly recommend to follow the given instructions in order to obtain accurate and reliable results.

The conformity of the **HCV REAL-TIME PCR Kit** to the prescribed technical requirements is subject to compliance of storage, transportation and handling conditions recommended by manufacturer.

6. WARNINGS AND PRECAUTIONS

Only personnel trained in the methods of molecular research are allowed to work with the kit.

Handle and dispose all biological samples, reagents and materials used to carry out the analysis as if they were able to transmit infective agents. The samples must be exclusively employed for certain type of analysis. Samples must be handled under a laminar flow hood. Tubes containing different samples must never be opened at the same time. Pipettes used to handle samples must be exclusively employed for this specific purpose. The pipettes must be of the positive dispensation type or be used with aerosol filter tips. The tips employed must be sterile, free from the DNases and RNases, free from DNA and RNA. The reagents must be handled under a laminar flow hood. The reagents required for amplification must be prepared in such a way that they can be used in a single session. Pipettes used to handle reagents must be exclusively employed for this specific purpose. The pipettes must be of the positive dispensation type or be used with aerosol filter tips. The tips employed must be sterile, free from the DNases and RNases, free from DNA and RNA. Avoid direct contact with the biological samples reagents and materials used to carry out the analysis. Wear powder-free surgical gloves. Wear protective clothing (work clothes and personal protective equipment) working with microorganisms classified as particularly pathogenic. The protective clothing and personal protective equipment must comply with the work to be performed and

health and safety requirements. Avoid producing spills or aerosol. Any material being exposed to biological samples must be treated for at least 30 minutes with disinfecting solution or autoclaved for 1 hour at 121 °C before disposal.

Molecular biology procedures, such as nucleic acids extraction, reverse transcription, PCR-amplification and detection require qualified staff to avoid the risk of erroneous results, especially due to the degradation of nucleic acids contained in the samples or sample contamination by amplification products.

All oligonucleotide components are produced by artificial synthesis technology according to internal quality control protocol and do not contain blood or products of blood processing.

Positive control is produced by artificial DNA synthesis technology. Positive control does not include parts of infectious agents.

All the liquid solutions are designed for single use and cannot be used more than once in amplification reactions. Plastic tubes do not contain phthalates. Do not breathe gas/fumes/vapor/spray produced by the components of the kit. Do not eat/drink components of the kit. Avoid contact with eyes. Only use the reagents provided in the kit and those recommended by manufacturer. Do not mix reagents from different batches. Do not use reagents from third party manufacturers' kits. All laboratory equipment, including pipettes, test tube racks, laboratory glassware, lab coats, bouffant caps, etc., as well as reagents should be strictly stationary. It is not allowed to move them from one room to another. Equip separate areas for the extraction/preparation of amplification reactions and for the amplification/detection of amplification products. Never introduce an amplification product in the area designed for extraction/preparation of amplification reactions. Wear lab coats, gloves and tools, which are exclusively employed for the extraction/preparation of the amplification reaction and for the amplification/detection of the amplification products. Never transfer lab coats, gloves and tools from the area designed for amplification/detection of the amplification products to the area designed for extraction/preparation of amplification reactions. Amplification products must be handled in such a way as to reduce dispersion into the environment as much as possible, in order to avoid the possibility of contamination. Pipettes used to handle amplification products must be exclusively employed for this specific purpose. Remove PCR waste only in a closed form. Remove waste materials (tubes, tips) only in a special closed container containing a disinfectant solution. Work surfaces, as well as rooms where NA extraction and PCR are performed, must be irradiated with bactericidal irradiators for 30 minutes before and after the work.

Do not open the tubes after amplification. Waste materials are disposed of in accordance with local and national standards. All surfaces in the laboratory (work tables, test tube racks, equipment, etc.) must be treated daily with disinfecting solution.

Emergency actions

Inhalation: Inhalation of the PCR-mix contained within this kit is unlikely, however care should be taken.

Eye Contact: If any component of this kit enters the eyes, wash eyes gently under potable running water for 15 minutes or longer, making sure that the eyelids are held open. If pain or irritation occurs, obtain medical attention.

Skin Contact: If any component of this kit contacts the skin and causes discomfort, remove any contaminated clothing. Wash affected area with plenty of soap and water. If pain or irritation occurs, obtain medical attention.

Ingestion: If any component of this kit is ingested, wash mouth out with water. If irritation or discomfort occurs, obtain medical attention.

Do not use the kit:

- When the transportation and storage conditions are breached;
- When the reagents' appearance does not respond to the kit passport;
- When the kit components packaging is breached;
- After the expiry date provided.

Significant health effects are NOT anticipated from routine use of this kit when adhering to the instructions listed in the current manual.

7. SAMPLES

Blood plasma/serum obtained from human peripheral whole blood is used for the analysis.

Interfering substances

The presence of PCR inhibitors in a sample may cause doubtful (unreliable) results. The sign of PCR inhibition is the simultaneous absence of internal control and specific product of amplification.

The maximum concentration of interfering substances, which does not affect the amplification of the laboratory control and internal control RNA-IC "A": triglycerides – up to 40 mmol/L of plasma/serum sample, hemoglobin – up to 2.0 g/L, bilirubin – up to 340 µmol/L, total protein - 80 g/L.

Concentrations of exogenous substances in biomaterial samples (blood plasma/serum) that do not affect the RT-PCR are as follows: peg-interferon alfa, acyclovir, atazanavir, ribavirin, rifampicin, isoniazid, azithromycin – up to three times the maximum therapeutic concentration.

General recommendations

The quality of sampling, sample storage, transport and pretreatment are of great importance for obtaining correct results.

Incorrect sampling may lead to unreliable results and, therefore, to the necessity for repeated sampling.

Sample collection

Whole peripheral blood must be collected in accordance with **PREP-NA-ULTRA**, **PREP-MB-ULTRA** and **PREP-MB-LITE** extraction kits' instructions for use.

Transportation and storage of the samples

Whole peripheral blood must be collected in accordance with **PREP-NA-ULTRA**, **PREP-MB-ULTRA** and **PREP-MB-LITE** extraction kits' instructions for use.

Sample preparation (obtaining blood plasma/serum)

Whole peripheral blood must be collected in accordance with **PREP-NA-ULTRA**, **PREP-MB-ULTRA** and **PREP-MB-LITE** extraction kits' instructions for use.

8. PROCEDURE

WARNING! The risk of cross-contamination between samples at all stages of the test, especially during aliquoting and RNA extraction, is a serious concern when performing research in a laboratory. Cross-contamination with high-copy biomaterial can lead to sporadic false-positive results.

To prevent cross-contamination of the biological material in the laboratory, the following rules are recommended:

1. conduct a visual assessment of the incoming biomaterial and dispose of the test tubes with broken integrity;
2. use negative control in each protocol, starting from RNA extraction stage;
3. use tips with aerosol filters at all stages of the analysis;
4. observe the analysis methods; open Eppendorf-type tubes without touching the inner part of the tube cap with gloved hand;
5. do not touch the edge of the tube with the tip when adding reagents.

8.1. RNA extraction

To extract HCV RNA from blood plasma/serum, reagent kits for viral NA extraction from blood are used (**PREP-NA-ULTRA**, **PREP-MB-ULTRA** and **PREP-MB-LITE** manufactured by "DNA-Technology R&P", LLC).

RNA extraction is performed according to the instructions to the used extraction kit.

WARNING! The obtained RNA preparation must be used within two hours for the reverse transcription reaction and the polymerase chain reaction.

8.2. The use of controls at the stage of nucleic acid extraction

8.2.1 Internal control RNA-IC "A"

To exclude false negative results and to control the quality of the test, an **internal control RNA-IC "A"** must be added into the test samples during NA extraction.

If using **PREP-NA-ULTRA** extraction kit, add 10 µL of RNA-IC "A" per sample.

If using **PREP-MB-ULTRA** and **PREP-MB-LITE** extraction kits, add 20 µL of RNA-IC "A" per sample.

8.2.2 Negative control (from PREP-NA-ULTRA, PREP-MB-ULTRA and PREP-MB-LITE extraction kits)

To exclude false positive results and to control the quality of the detection, a **negative control must be used**.

At the stage of nucleic acid extraction, a negative control (negative control from the **PREP-NA-ULTRA**, **PREP-MB-ULTRA** extraction kits or dilution solution for controls from the **PREP-MB-LITE** extraction kit) must be prepared and run through all stages of extraction simultaneously with the RNA extraction from test samples in accordance with the instructions for use of the corresponding reagent kit. Each group of extracted samples must include one negative control ("C-").

8.3. Reverse transcription and polymerase chain reaction (RT-PCR)

WARNING!

1. The reagents and tubes should be kept away from direct sun light.
2. Strictly observe the completeness of the strips and caps for them. Do not use the caps for the strips of the other kits!

8.3.1. Mark one tube/stripped tube with the paraffin sealed PCR-mix for each test sample, negative control (C-), positive control (C+).

WARNING! The reagents are intended for up to 6 runs, considering variable number of test samples, 1 negative control and 1 positive control per run).

Example: to test 6 samples, mark 6 tubes (one for each sample), one for "C-" and one for "C+"). The total number of tubes is 8.

8.3.2. Shake the tubes with RT-PCR-buffer "V" and Enzyme Taq/RT on vortex mixer for 1-3 seconds and spin in a vortex mixer for 3-5 seconds.

WARNING! Take Enzyme Taq/RT out of the freezer immediately prior to use.

8.3.3. Prepare the mixture of RT-PCR-buffer "V" and Enzyme Taq/RT. Add to the one tube:

- 15.0 x (N+1) µL of RT-PCR-buffer "V";
- 0.5 x (N+1) µL of Enzyme Taq/RT,

N is the number of test samples considering "C-", "C+".

The mixture can be stored at temperature from 2 °C to 8 °C for no longer than one hour.

Example: to test 6 samples (8 marked tubes), prepare the mixture of RT-PCR-buffer "V" and Enzyme Taq/RT for 9 (8+1) tubes, i.e. mix 135 µL of RT-PCR-buffer "V" and 4.5 µL of Enzyme Taq/RT.

WARNING! When taking the Enzyme Taq/RT, do not dip the tip deeper than 1.0 mm and observe the rules for viscous liquids dosing. Thoroughly flush the remaining Enzyme Taq/RT from the tip by pipetting at least 5 times.

8.3.4. Shake the tube with the mixture of RT-PCR-buffer “V” and Enzyme Taq/RT on vortex and spin on vortex for 1-3 seconds.

8.3.5. Add 15 µL of the RT-PCR-buffer “V” and Enzyme Taq/RT mixture into each tube, including “C-” and “C+”. Avoid paraffin layer break. Cover the tubes/strips loosely with caps.

WARNING! After adding the mixture of RT-PCR-buffer “V” and Enzyme Taq/RT into the tubes with PCR-mix, immediately proceed to 8.3.6 – 8.3.11.

8.3.6. Shake the tube with positive control “C+” on vortex for 3-5 seconds and spin on vortex for 1-3 seconds.

WARNING!

1. Before introducing RNA preparation and negative control into tubes with PCR-mix, fulfill the recommendations for RNA preparation use listed in the **PREP-NA-ULTRA**, **PREP-MB-ULTRA** and **PREP-MB-LITE** reagent kits instructions.

2. To prevent contamination, only open the caps of the tubes into which the RNA sample is to be added and close them before adding the next RNA sample. If strips are used, close the strip caps after adding the sample before proceeding with the next sample. Close the tubes/strips tightly. Use filter tips.

8.3.7. Add 20 µL of RNA sample into the corresponding marked tubes. Do not add RNA into “C-”, “C+” tubes. Avoid paraffin layer break.

8.3.8. Add 20 µL of negative control (“C-”), which passed whole RNA extraction procedures into the corresponding tube.

8.3.9. Add 20 µL of positive control sample (“C+”) into the corresponding tube. Avoid paraffin layer break.

8.3.10. Spin the tubes on vortex for 3–5 seconds.

8.3.11. Set the tubes/strips into the real-time thermal cycler and run RT-PCR.

8.3.12. **For DT detecting thermal cyclers:**

Launch the operating software for DT instrument². Add corresponding test³, specify the number and IDs of the samples, positive and negative control samples. Specify the position of the tubes/strips in the thermal unit (see 8.3.11) and run RT-PCR. See Table 3.

8.3.13. **For CFX96 and Applied Biosystems QuantStudio 5 detecting thermal cyclers:**

Run RT-PCR considering the PCR-mix volume of 50 µL. See Tables 4, 5.

² Please, apply to Operation Manual for DTprime and DTlite Real-Time PCR instruments PART II.

³ Instructions for uploading "files with test parameters" can be found on "DNA-Technology's" website <https://www.dna-technology.com/assaylibrary>.

Table 3. The RT-PCR program for DTlite and DTprime

Step	Temperature, °C	Min.	Sec.	Number of cycles	Optical measurement	Type of the step
1	47	15	0	1		Cycle
2	95	5	0	1		Cycle
3	95	0	10	50		Cycle
	59	0	20		√	
5	25 ¹	...	...	Holding		Holding
√ - optical measurement ¹ – holding at 10 °C is allowed						

Table 4. The RT-PCR program for CFX96 (Bio-Rad)

Step	Temperature, °C	Time, min:sec	Cycle repeats
1	47	15:00	1
2	95	5:00	1
3	95	0:10	50
4	59 √	0:20	
√ - optical measurement (Plate Read), set the fluorescence measurement on the Fam and Hex channels at 59 °C			

Table 5. The RT-PCR program for Applied Biosystems QuantStudio 5*

Step	No.	Temperature, °C	Time min:sec	Number of cycles (repeats)
Holding	1	47	15:00	1
	2	95	05:00	1
PCR	1	95	0:10	50
	2	59 √	0:25	
√ - data collection for fluorophores (Fam, Vic (Hex)) is on				
* - experiment type is Standard curve. Quick Run mode is acceptable.				

9. CONTROLS

The **HCV REAL-TIME PCR Kit** contains positive control. Positive control is a cloned part of the virus genome. It is produced with genetic engineering techniques and characterized by automatic sequencing. The kit includes the Internal control RNA-IC "A". RNA-IC "A" is intended to assess the quality of the RNA extraction and polymerase chain reaction. To reveal possible contamination a negative control is required.

WARNING! A negative control sample should go through all stages of RNA extraction. Negative control from an extraction kit can be used as a negative control in volumes indicated in supplied instructions.

The test result is considered valid when:

- the exponential growth of the fluorescence level for the specific product is present, in this case the internal control is not considered;
- the exponential growth of the fluorescence level for the specific product is absent and for internal control is present.

The test result is considered invalid when the exponential growth of the fluorescence level for the specific product and for internal control is not observed.

If positive control (C+) does **not** express growing fluorescence of the specific product or positive result, it is required to repeat the whole test. It may be caused by inhibitors, operation error or violation of storage and handling.

If negative control (C-) expresses growing fluorescence of the specific product or positive result, all tests of the current batch are considered false. Decontamination is required.

10. DATA ANALYSIS

- 10.1. The Real-time PCR thermal cycler software delivered with the thermal cycler analyzes the results automatically.
- 10.2. The interpretation should be performed in accordance with Table 6. Run results are valid if the result interpretation conditions for controls are observed.
- 10.3. When using CFX96 detection thermal cyclers, use regression type analysis (Cq Determination Mode: Regression) and exclude the first 5 cycles from the analysis (Analyze Date from Cycle 5 to 50).
- 10.4. When using the Applied Biosystems QuantStudio 5 detecting thermal cycler, amplification data can be obtained by different methods, e.g. base threshold (Ct) or relative threshold (Crt). Therefore, these points may have different abbreviations (Ct, Crt) but are further processed in the same way. In the relative threshold (Crt) setting, the initial cycle of Crt is "5". The interpretation of the RNA-IC "A" amplification results (Vic) in Table 6 corresponds to the relative threshold (Crt).
- 10.5. When using CFX96 and Applied Biosystems QuantStudio 5 detecting thermal cyclers, analyze the selected samples and exclude the wells which signal exceeds the background signal of the instrument, but has a linear, not S-shaped character.

Table 6. RT-PCR results interpretation

Detection channel		Result interpretation
Fam (HCV RNA) Cp/Cq/Crt	Hex/Vic (IC) Cp/Cq/Crt	
Test samples		
Specified	Not considered	HCV RNA is detected
Not specified	≤35	HCV RNA is not detected
Not specified	>35 or not specified	Unreliable result
Negative control		
Not specified	≤35	Negative result Run results are valid
Positive control		
Specified	Not considered	Positive result Run results are valid

- 10.6. Unreliable results may be caused by the presence of inhibitors in the nucleic acid preparation obtained from the biological material, errors in the pre-analytical stage, incorrect implementation of the analysis Protocol, non-compliance with the temperature mode of amplification, etc. In this case, either repeated RT-PCR with the RNA preparation, or repeated RNA extraction and RT-PCR, or repeated biomaterial collection (performed sequentially) is required.
- 10.7. If a negative result is obtained for a positive control, the results of the whole run batch are considered invalid. In this case the repeated run for the whole batch of samples is required.
- 10.8. If a positive result for a negative control, the results of the entire run are considered unreliable. In this case, special measures are required to detect and eliminate possible contamination.

11. SPECIFICATIONS

a. Analytical specificity

In samples of human biological material containing hepatitis C virus RNA, the amplification software of the detecting thermal cycler records a positive amplification result for the specific product on the Fam detection channel.

The reagent kit detects the following HCV genotypes: 1a, 1b, 2, 3, 4, 5, 6.

In samples of biological material not containing HCV RNA, the amplification software of the detecting thermal cycler records a negative result of specific product amplification on the Fam detection channel and a positive result of internal control (IC) amplification by the Hex detection channel.

No nonspecific positive amplification results were registered for the NA of the following viruses: RNA of HAV, HDV, HGV, HIV; DNA of HBV, EBV, CMV, and human DNA at up to 1.0×10^8 copies/mL of the sample. No cross-reactions for the indicated microorganisms and viruses were registered.

b. Analytical sensitivity (limit of detection)

Limit of detection (if using **PREP-NA-ULTRA**, **PREP-MB-ULTRA** and **PREP-MB-LITE** extraction kits):

- 10 IU/mL for RNA extraction from **1,000 µL** of blood plasma/serum using the **PREP-NA-ULTRA** reagent kit;
- 15 IU/mL for RNA extraction from **500 µL** of blood plasma/serum using the **PREP-NA-ULTRA** reagent kit;
- 50 IU/mL for RNA extraction from **250 µL** of blood plasma/serum using the **PREP-NA-ULTRA** reagent kit;
- 15 IU/mL for RNA extraction from **500 µL** of blood plasma/serum using the **PREP-MB-ULTRA** reagent kit;
- 50 IU/mL for RNA extraction from **250 µL** of blood plasma/serum using the **PREP-MB-LITE** reagent kit;
- 80 IU/mL for RNA extraction from **100 µL** of blood plasma/serum using the **PREP-MB-LITE** reagent kit.

The limit of detection was established by analyzing serial dilutions of two series of laboratory controls.

c. Diagnostic characteristics

Number of samples (n) – 93;

Diagnostic sensitivity (95% CI) – 100% (91.96-100%);

Diagnostic specificity (95% CI) – 100% (92.75-100%).

12. TROUBLESHOOTING

Table 7. Troubleshooting

	Result	Possible cause	Solution
C+	-	Operation error PCR inhibition Violation of storage and handling requirements	Repeat the whole test Dispose of the current batch
C-	+	Contamination	Dispose of the current batch Perform decontamination procedures
IC	-	PCR inhibition RNA extraction violation	Repeat RNA extraction Repeat the whole test Resample

If you face to any undescribed issues contact our customer service department:

Phone: +7(495)640.16.93

E-mail: hotline@dna-technology.ru

<https://www.dna-technology.com/support>

13. QUALITY CONTROL

The quality control procedures performed in accordance with ISO 9001:2015 and ISO 13485:2016:

- observation of quality management in manufacturing of products;
- creation of values for customers;
- maintenance of the best service quality and customer management.

Technical support:

E-mail: hotline@dna-technology.ru

<https://www.dna-technology.com>

Manufacturer:

"DNA-Technology Research & Production", LLC,

142281, Russia, Moscow Region,

Protvino, Zheleznodorozhnaya Street, 20

Phone/fax: +7(495) 640.17.71

E-mail: info@dna-technology.com

<https://www.dna-technology.com>

Seller: "DNA-Technology" LLC,

117587, Russia, Moscow,

int. ter. Municipal District Chertanovo Severnoye,

Varshavskoye shosse, 125 Zh, building 5, floor 1, office 12;

Phone/fax: +7(495) 640.17.71

E-mail: info@dna-technology.com

<https://www.dna-technology.com>

14. KEY TO SYMBOLS

	For research use only		Date of manufacture
	Temperature limit		Consult instructions for use
	Contains sufficient for <n> tests		Catalogue number
	Use-by date		Manufacturer
	Batch code		Keep away from sunlight
	Version		Non-sterile
	Caution		

REF

R3-P613-S3/9ER
R3-P613-23/9ER

VER

723-3.2025.04.07