



1534 2026-06-05

**For professional use only**

Helicobacter pylori REAL-TIME PCR Detection Kit

INSTRUCTION FOR USE



"DNA-Technology Research & Production", LLC,

142281, Russia,

Moscow Region, Protvino,

Zheleznodorozhnaya Street, 20

Phone/fax: +7(495) 640.17.71

E-mail: info@dna-technology.com

<https://www.dna-technology.com>

Customer service department

E-mail: hotline@dna-technology.ru



R1-P501-S3/9EU
R1-P501-23/9EU
R1-P501-UA/9EU



1534.2026.06.05

TABLE OF CONTENTS

1. INTENDED USE	3
2. METHOD	3
3. CONTENT	4
4. REAGENTS AND EQUIPMENT REQUIRED BUT NOT PROVIDED	4
5. TRANSPORT AND STORAGE CONDITIONS	6
6. WARNINGS AND PRECAUTIONS	7
7. SAMPLES	8
8. PROCEDURE	10
9. CONTROLS	16
10. DATA ANALYSIS	16
11. SPECIFICATIONS	17
12. TROUBLESHOOTING	18
13. QUALITY CONTROL	18
14. KEY TO SYMBOLS	20

1. INTENDED USE

The **Helicobacter pylori REAL-TIME PCR Detection Kit** is an *in vitro* Nucleic Acid Test (NAT) – pathogen-detection-based product. The **Helicobacter pylori REAL-TIME PCR Detection Kit** is designed to detect *Helicobacter pylori* DNA in human biological samples with an aid of Polymerase Chain Reaction (PCR) method. Samples are human biological materials: biotates, faeces.

Indications for the use: symptoms of infectious or inflammatory process in the gastrointestinal tract, control of treatment against *Helicobacter pylori*.

The application of the kit does not depend on population and demographic aspects. There are no contradictions for use of the **Helicobacter pylori REAL-TIME PCR Detection Kit**.

The **Helicobacter pylori REAL-TIME PCR Detection Kit** can be used in clinical and diagnostic laboratories of medical institutions and research practice.

Potential users: personnel qualified in molecular diagnostics methods and working in the clinical and diagnostic laboratory.

It is necessary to apply the kit only as directed in this instruction for use.

2. METHOD

The implemented PCR method is based on amplification of a target DNA sequence. To increase the sensitivity and specificity of the amplification reaction, the use of a hot-start is provided. Hot-start is provided by reaction mixture preparation consisting of two layers separated by a layer of paraffin or the use of Taq polymerase blocked by antibodies. The polymerase chain reaction starts only when paraffin is melted or thermal dissociation of a complex of Taq polymerase and antibodies is happened. It excludes non-specific annealing of primers to targets DNA in the initial heating of the tube.

The **Helicobacter pylori REAL-TIME PCR Detection Kit** is based on fluorescent modification of the PCR method. The PCR mix contains two target-specific probes bearing reporter fluorescent dyes (Fam and Hex) and quencher molecules. Once hybridized to a target sequence, the probes become activated. As a result of activation fluorescence increases proportionally to target sequence amplification. The intensity of fluorescence is measured at every cycle of reaction with a Real-time PCR thermal cycler data collection unit and analyzed with the software provided.

The PCR mix includes the Internal control (IC), which is intended to assess the quality of the polymerase chain reaction. DNA probes used for the detection of the *Helicobacter pylori* product amplification includes fluorescent dyes Fam. DNA probe used for the detection of the internal control amplification product includes the fluorescent dye Hex. Table 1 shows the detection channels of amplification products.

Table 1. Detection channels of amplification products

Fam/Green	Hex/Yellow	Rox/Orange	Cy5/Red	Cy5.5/Crimson
<i>Helicobacter pylori</i>	IC	-	-	-

The automatic analysis is available on “DNA-Technology” made instruments DTlite or DTprime REAL-TIME Thermal Cyclers or Bio-Rad made iCycler iQ or iQ5 or QIAGEN made Rotor-Gene Q for **Helicobacter pylori REAL-TIME PCR Detection Kit** (see the catalogue at <https://www.dna-technology.com> to see available supply options). The current version of the software is available for download at <https://www.dna-technology.com/software>.

3. CONTENT

The detailed description of content is represented in Tables 2-4.

Table 2. The **Helicobacter pylori REAL-TIME PCR Detection Kit** content, package S (standard), strips for R1-P501-S3/9EU

Reagent	Description	Volume per tube	Amount
Paraffin-sealed PCR mix	Colorless transparent liquid under waxy white fraction	20 µL in each	Tubes, 12 strips of 8
Taq polymerase solution	Colorless transparent liquid	500 µL in each	2 tubes
Mineral oil	Colorless transparent viscous oily liquid	1.0 mL in each	2 tubes
Positive control	Colorless transparent liquid	130 µL	1 tube
Strip caps	12 strips of 8		

Table 3. The **Helicobacter pylori REAL-TIME PCR Detection Kit** content, package S (standard), tubes for R1-P501-23/9EU

Reagent	Description	Volume per tube	Amount
Paraffin-sealed PCR mix	Colorless transparent liquid under waxy white fraction	20 µL in each	96 individual tubes
Taq polymerase solution	Colorless transparent liquid	500 µL in each	2 tubes
Mineral oil	Colorless transparent viscous oily liquid	1.0 mL in each	2 tubes
Positive control	Colorless transparent liquid	130 µL	1 tube

Table 4. The **Helicobacter pylori REAL-TIME PCR Detection Kit** content, package U, for R1-P501-UA/9EU

Reagent	Description	Volume per tube	Amount
PCR mix	Colorless or slightly pink transparent liquid	600 µL	1 tube
TechnoTaq MAX polymerase	Colorless transparent viscous liquid	30 µL	1 tube
PCR buffer	Colorless transparent liquid	600 µL	1 tube
Positive control	Colorless transparent liquid	130 µL	1 tube

All components are ready to use and do not require additional preparation for operation.

The **Helicobacter pylori REAL-TIME PCR Detection Kit** is intended for single use and designed for 96 tests (no more than 94 defined samples, one positive control and one negative control).



The kit in the package U is intended for 96 samples and requires no less than 5 samples in a single run (3 test samples, positive and negative controls).

4. REAGENTS AND EQUIPMENT REQUIRED BUT NOT PROVIDED

4.1. Specimen collection

- Sterile containers to collect clinical material;
- Sterile tubes containing transport media: “DNA-Technology” **STOR-F** (**REF** P-901-1/1EU, P-901-N/1EU, P-901-R/1EU) or equivalent or sterile physiological saline solution for the transportation of the sample.

4.2. DNA extraction and PCR

Preamplification-specimen and control preparation area:

- Biological safety cabinet class II;
- Refrigerator;
- Vortex mixer;
- High speed centrifuge (RCF(g) no less than 16000);
- Solid-state thermostat (temperature range 40-95 °C);
- Nucleic acid extraction kit (“DNA-Technology” made **PREP-NA** (**REF** P-002/1EU) or **PREP-GS** (**REF** P-003/1EU) extraction kits are recommended);
- Tube rack for 1.5 mL tubes;
- 1.5 mL tubes;
- Physiological saline solution 0.9% NaCl (Sterile);
- Electric laboratory aspirator with trap flask for the removal of supernatant;
- Single channel pipettes (dispensers covering 20-1000 µL volume range);
- RNase and DNase free filtered pipette tips (volume 200 µL, 1000 µL);
- RNase and DNase free pipette tips for aspirator with trap flask;
- Container for used pipette tips, tubes and other consumables;
- Powder-free surgical gloves;
- Disinfectant solution.

Preamplification-reagent preparation area:

- UV PCR cabinet;
- Refrigerator;
- Freezing chamber (using detection kit in the package U **REF** R1-P501-UA/9EU);
- Vortex mixer;
- Vortex rotor for strips (using detection kit package in strips **REF** R1-P501-S3/9EU);
- Tube rack for 1.5 mL tubes;
- PCR tube rack for 0.2 mL tubes or strips;
- 0.2 mL PCR tubes (using detection kit in the package U **REF** R1-P501-UA/9EU);
- Single channel pipettes (dispensers covering 2.0-1000 µL volume range);
- RNase and DNase free filtered pipette tips (volume 20 µL, 200 µL, 1000 µL);
- DTstream M1 dosage instrument (only for automated dosing using detection kit in the package U **REF** R1-P501-UA/9EU);
- Device for tray sealing DTpack (“DNA-Technology”, LLC) (only for automated dosing using detection kit in the package U **REF** R1-P501-UA/9EU);
- Centrifuge for microtrays (only for automated dosing using detection kit in the package U **REF** R1-P501-UA/9EU);

- Polymer thermal seal for microtray sealing (only for automated dosing using detection kit in the package U **REF** R1-P501-UA/9EU);
- PCR microtray (only for automated dosing using detection kit in the package U **REF** R1-P501-UA/9EU);
- Container for used pipette tips, tubes and other consumables;
- Powder-free surgical gloves;
- Disinfectant solution.

Post-Amplification – Amplification detection area:

- Real-time PCR thermal cycler.

Software:

The most recent version of the DT thermal cyclers software can be downloaded from <https://www.dna-technology.com/software>.

The OS supported: all versions of Windows starting from 7.

5. TRANSPORT AND STORAGE CONDITIONS

Expiry date – 12 months from the date of production.

All components of the **Helicobacter pylori REAL-TIME PCR Detection Kit** except TechnoTaq MAX polymerase must be stored at temperatures from 2 °C to 8 °C during the storage period. TechnoTaq MAX polymerase must be stored at temperatures from minus 18 °C to minus 22 °C during the storage period. PCR mix must be stored at temperatures from 2 °C to 8 °C and out of light during the storage period. The excessive temperature and light can be detrimental to product performance.

Transportation of the kit except TechnoTaq MAX polymerase is allowed in thermal containers with icepacks by all types of covered transport at temperatures from 2 °C to 25 °C inside the container, but for no longer than 5 days.

It is allowed to transport TechnoTaq MAX polymerase (package U) in thermal containers with icepacks at temperatures from 2 °C to 8 °C inside the container, but for no longer than 5 days.

Shelf-life of the kit following the first opening of the primary container:

- components of the kit except TechnoTaq MAX polymerase should be stored at temperatures from 2 °C to 8 °C during the storage period;
- TechnoTaq MAX polymerase should be stored at temperatures from minus 18 °C to minus 22 °C during the storage period;
- PCR mix for amplification should be stored at temperatures from 2 °C to 8 °C and out of light during the storage period.

The kit stored in under undue regime should not be used.

An expired the **Helicobacter pylori REAL-TIME PCR Detection Kit** should not be used.

We strongly recommend to follow the given instructions in order to obtain accurate and reliable results.

The conformity of the **Helicobacter pylori REAL-TIME PCR Detection Kit** to the prescribed technical requirements is subject to compliance of storage, transportation and handling conditions recommended by manufacturer.

Contact our official representative in EU by quality issues of the **Helicobacter pylori REAL-TIME PCR Detection Kit**.

6. WARNINGS AND PRECAUTIONS

Only personnel trained in the methods of molecular diagnostics and the rules of work in the clinical and diagnostic laboratory are allowed to work with the kit.

Handle and dispose all biological samples, reagents and materials used to carry out the assay as if they were able to transmit infective agents. The samples must be exclusively employed for certain type of analysis. Samples must be handled under a laminar flow hood. Tubes containing different samples must never be opened at the same time. Pipettes used to handle samples must be exclusively employed for this specific purpose. The pipettes must be of the positive dispensation type or be used with aerosol filter tips. The tips employed must be sterile, free from the DNases and RNases, free from DNA and RNA. The reagents must be handled under a laminar flow hood. The reagents required for amplification must be prepared in such a way that they can be used in a single session. Pipettes used to handle reagents must be exclusively employed for this specific purpose. The pipettes must be of the positive dispensation type or be used with aerosol filter tips. The tips employed must be sterile, free from the DNases and RNases, free from DNA and RNA. Avoid direct contact with the biological samples reagents and materials used to carry out the assay. Wear powder-free surgical gloves. Wear protective clothing (work clothes and personal protective equipment) working with microorganisms classified as particularly pathogenic. The protective clothing and personal protective equipment must comply with the work to be performed and health and safety requirements. Avoid producing spills or aerosol. Any material being exposed to biological samples must be treated for at least 30 minutes with disinfecting solution or autoclaved for 1 hour at 121 °C before disposal.

Molecular biology procedures, such as nucleic acids extraction, PCR-amplification and detection require qualified staff to avoid the risk of erroneous results, especially due to the degradation of nucleic acids contained in the samples or sample contamination by amplification products.

All oligonucleotide components are produced by artificial synthesis technology according to internal quality control protocol and do not contain blood or products of blood processing.

Positive control is produced by artificial DNA synthesis technology. Positive control does not include parts of infectious agents.

All the liquid solutions are designed for single use and can not be used more than once in amplification reactions. Plastic tubes do not contain phthalates. Do not breathe gas/fumes/vapor/spray produced by the components of the kit. Do not eat/drink components of the kit. Avoid contact with eyes. Only use the reagents provided in the kit and those recommended by manufacturer. Do not mix reagents from different batches. Do not use reagents from third party manufacturers' kits. All laboratory equipment, including pipettes, test tube racks, laboratory glassware, lab coats, bouffant caps, etc., as well as reagents should be strictly stationary. It is not allowed to move them from one room to another. Equip separate areas for the extraction/preparation of amplification reactions and for the amplification/detection of amplification products. Never introduce an amplification product in the area designed for extraction/preparation of amplification reactions. Wear lab coats, gloves and tools, which are exclusively employed for the extraction/preparation of the amplification reaction and for the amplification/detection of the amplification products. Never transfer lab coats, gloves and tools from the area designed for amplification/detection of the amplification products to the area designed for extraction/preparation of amplification reactions. Amplification products must be handled in such a way as to reduce dispersion into the environment as much as possible, in order to avoid the possibility of contamination. Pipettes used to handle amplification products must be exclusively employed for this specific purpose. Remove PCR waste only in a closed form. Remove waste materials (tubes, tips) only in a special closed container containing a disinfectant solution. Work surfaces, as well as rooms where NA extraction and PCR are performed, must be irradiated with bactericidal irradiators for 30 minutes before and after the work.

Do not open the tubes after amplification. Waste materials are disposed of in accordance with local and national standards. All surfaces in the laboratory (work tables, test tube racks, equipment, etc.) must be treated daily with disinfecting solution.

Emergency actions

Inhalation: Inhalation of the PCR mix contained within this kit is unlikely, however care should be taken.

Eye Contact: If any component of this kit enters the eyes, wash eyes gently under potable running water for 15 minutes or longer, making sure that the eyelids are held open. If pain or irritation occurs, obtain medical attention.

Skin Contact: If any component of this kit contacts the skin and causes discomfort, remove any contaminated clothing. Wash affected area with plenty of soap and water. If pain or irritation occurs, obtain medical attention.

Ingestion: If any component of this kit is ingested, wash mouth out with water. If irritation or discomfort occurs, obtain medical attention.

Do not use the kit:

- When the transportation and storage conditions are breached;
- When the reagents' appearance does not respond to the kit passport;
- When the kit components packaging is breached;
- After the expiry date provided.

Significant health effects are **NOT** anticipated from routine use of this kit when adhering to the instructions listed in the current manual.

7. SAMPLES

The **Helicobacter pylori REAL-TIME PCR Detection Kit** is designed to detect DNA extracted from the bioplates and faeces depending on professional prescription.

Sampling, sample processing procedures and storage are carried out in accordance with the instructions to the DNA extraction kit from biological material.

Interfering substances

The presence of PCR inhibitors in a sample may cause controversial (uncertain) results. The sign of PCR inhibition is the simultaneous absence of internal control and specific product of amplification.

PCR inhibitors are the presence of hemoglobin in a DNA sample as a result of incomplete removal during DNA extraction from biomaterial sample containing blood impurities, as well as the presence of isopropyl alcohol and methyl acetate in a DNA sample as a result of incomplete removal of washing solutions during sample preparation.

The maximum concentrations of interfering substances, that have no effect on the amplification of the laboratory control sample and internal control are: hemoglobin – 0.35 mg/mL of the DNA sample, isopropyl alcohol – 100 µL/mL of the DNA sample, methyl acetate – 100 µL/mL of the DNA sample.

Impurities contained in the biomaterial sample are almost completely removed during the DNA extraction. To reduce the count of PCR inhibitors, it is necessary to follow the principles of taking biological material. Suspecting a large count of PCR inhibitors in the sample, it is recommended to choose DNA extraction methods that allow to remove PCR inhibitors from the sample as much as possible. It is not recommended to use express methods of DNA extraction.

General requirements

To interpret results successfully and robustly, a high quality of sample and appropriate conditions of storage, transport, and handling are required.

PCR analysis refers to direct methods of laboratory research; therefore the collection of biological material must be carried out from the site of infection localization.

Professional prescription is required to localize the place of sampling. The decision must be based on a patient's complaints and clinical signs, and made by the physician in charge.

Sample collection

Bioptates sampling

Bioptates are transferred to a 1.5 mL tubes with transport medium intended by the manufacturer for transportation and storage of samples for PCR. After sample collection the tube is tightly closed and marked.

Faeces sampling

Samples of faeces with mass (volume) 250 mg (250 µL) are transferred to a 1.5 tubes with 1.0 mL of sterile physiological saline. After sample collection the tube is tightly closed and marked.

Transportation and storage of the samples



Samples are stored according to the instruction for the transport medium used intended for subsequent sample analysis by PCR.

Samples may be transported and stored in physiological saline at temperatures from 2 °C to 8 °C for no more than 24 hours. When it is impossible to deliver the material in the laboratory during the day, a one-time freezing of the material is allowed. The frozen material is allowed to be stored at temperatures from minus 18 °C to minus 22 °C for one month.

Sample preparation

Preparation of the bioptates:



For DNA extraction from bioptates only **PREP-NA** and **PREP-GS** are allowed.

It is necessary to perform pretreatment before DNA extraction by the in **PREP-NA** and **PREP-GS** kits.

1. Vortex the tubes with samples for 3-5 seconds and spin down drops for 3-5 seconds.
2. Remove the supernatant.

The following sample preparation is made according to user manual for the extraction kit used.

Preparation of the faeces:



For DNA extraction from faeces only **PREP-NA** is allowed.

It is necessary to perform pretreatment before DNA extraction by the in **PREP-NA** kit.

1. Vortex the tube with faeces in 1.0 mL of saline during 5-10 seconds.
2. Centrifuge the tube at RCF(g) 100 for 2-3 minutes.
3. Transfer 800–1000 µL liquid material to 1.5 mL plastic tube, centrifuge the tube at RCF(g) 16000 for 10 minutes.
4. Remove the supernatant, leaving 100 µL in tube (precipitate + liquid fraction). Tightly close the tubes.

The samples are ready for DNA extraction.



The detailed description of sampling and sample processing procedures as well as sample storage and transportation requirements are cited in **PREP-NA** and **PREP-GS** extraction kits user manuals.

8. PROCEDURE

DNA extraction from biological material

DNA extraction is carried out according to the extraction kit instructions. **PREP-NA** and **PREP-GS** extraction kits are recommended.



Independently of DNA extraction kit used, a negative control sample should go through all stages of DNA extraction. Physiological saline solution or negative control sample from an extraction kit can be used as a negative control sample in volumes as indicated.

Assay procedure

8.1 Preparing PCR for package S



The reagents and tubes should be kept away from direct sun light.



When using package S, strips (R1-P501-S3/9EU), strictly observe the completeness of the strips and caps for them. Do not use the caps for the strips of the other kits!

8.1.1 Mark the required number of tubes with paraffin-sealed PCR mix for each test sample, positive control (C+) and negative control (C-).

Example: to test 5 samples, mark 5 tubes for samples, 1 tube for “C-” and 1 tube for “C+”. The resulting number of tubes is 7.

8.1.2 Vortex the Taq polymerase solution for 3-5 seconds, then spin for 1-3 seconds to collect the drops.

8.1.3 Add 10 µL of Taq polymerase solution into each tube. Avoid paraffin layer break.

8.1.4 Add one drop (~20 µL) of mineral oil into each tube (not applicable to kits approved for use with Rotor-Gene Q thermal cycler).

8.1.5 Vortex the tubes with samples, “C+” and “C-” for 3-5 seconds and spin down the drops for 1-3 seconds.



In case of using **PREP-GS DNA Extraction Kit**. After vortexing centrifuge the tubes with the DNA preparation at RCF(g) 16000 for one minute to precipitate the sorbent. If, after isolation, the supernatant containing the isolated DNA was transferred to new tubes, centrifugation is carried out for 1-3 seconds in a vortex mixer.



Open the cap of the tube, add DNA sample (or control sample), then close the tube before proceeding to the next DNA sample to prevent contamination. In case of using tubes in strips, close the strip before proceeding to the next strip to prevent contamination. Close the tubes/strips tightly. Use filter tips.

8.1.6 Add 5.0 µL of DNA sample into corresponding tubes. Do not add DNA into the “C+”, “C-” tubes. Avoid paraffin layer break.

8.1.7 Add 5.0 µL of negative control (C-) which passed whole DNA extraction procedure into corresponding tube. Add 5.0 µL of positive control sample (C+) into corresponding tube. Avoid paraffin layer break.

8.1.8 Spin tubes/strips for 3-5 seconds (when using the Rotor-Gene Q thermal cycler, centrifugation is not required).

8.1.9 Set the tubes/strips into the Real-time Thermal Cycler.

8.1.10 Launch the operating software for DT instrument¹. Add corresponding test², specify the number and ID's of the samples, positive and negative control samples. Specify the position of the tubes/strips in the thermal unit (see 8.1.9) and run PCR. See Tables 5-9.

Table 5. The PCR program for DTLite and DTprime Thermal Cyclers

Step	Temperature, °C	Min.	Sec.	Number of cycles	Optical measurement	Type of the step
1	80	0	30	1		Cycle
	94	1	30			
2	94	0	30	5		Cycle
	64	0	15		v	
3	94	0	10	45		Cycle
	64	0	15		v	
4	94	0	5	1		Cycle
5	10 ¹	...	...	Holding		Holding
v - optical measurement						
¹ – holding at 25°C is allowed						

Table 6. The PCR program for iCycler iQ thermal cycler (with persistent well factor)

Cycle	Repeats	Step	Dwell time	Setpoint, °C	PCR/Melt Data Acquisition
1	1				
		1	1 min	80	
		2	1 min 30 sec	94	
2	5				
		1	30 sec	94	
		2	45 sec	64	
3	45				
		1	10 sec	94	
		2	45 sec	64	Real Time
4	...	...	...	10	Storage

¹ Please, apply to Operation Manual for DTprime and DTLite Real-Time PCR instruments PART II.

² Instructions for uploading "files with test parameters" can be found on "DNA-Technology's" website <https://www.dna-technology.com/assaylibrary>.

Table 7. The PCR program for iCycler iQ thermal cycler (with dynamic well factor)

Cycle	Repeats	Step	Dwell time	Setpoint, °C	PCR/Melt Data Acquisition
dynamicwf.tmo program					
1	1				
		1	1 min	80	
		2	1 min 30 sec	94	
2	5				
		1	30 sec	94	
		2	45 sec	64	
3	2				
		1	30 sec	80	Real Time
PCR program					
4	45				
		1	10 sec	94	
		2	45 sec	64	Real Time
5	...	...	...	10	Storage

Table 8. The PCR program for Rotor-Gene Q thermal cycler

Cycling	Temperature	Hold time	Cycle repeats
Cycling	80 deg	60 sec	1 time
	94 deg	90 sec	
Cycling 2	94 deg	30 sec	5 times
	57 deg*	15 sec	
Cycling 3	94 deg	10 sec	45 times
	57 deg*	15 sec	
* Take the measurement			

Table 9. Detection channels

Fam (Green)	Hex (Yellow)	Rox (Orange)	Cy5 (Red)	Cy5.5 (Crimson)
Specific product and C+	IC	-	-	-

8.2 Preparing PCR for package U



The reagents and tubes should be kept away from direct sun light.

8.2.1 Mark the required number of 0.2 mL tubes for each test sample, positive control (C+) and negative control (C-).

Example: to test 5 samples, mark 5 tubes for samples, 1 tube for "C-" and 1 tube for "C+". The resulting number of tubes is 7.

8.2.2 Vortex the tube with PCR mix for 3-5 seconds, then spin on vortex for 1-3 seconds to collect the drops.

8.2.3 Add to each tube 6.0 µL of PCR mix.

8.2.4 Vortex the TechnoTaq MAX polymerase and PCR buffer for 3-5 seconds, then spin for 1-3 seconds.



TechnoTaq MAX polymerase should be got out from the freezer immediately prior to use.

8.2.5 Prepare the mixture of PCR buffer and TechnoTaq MAX polymerase. Add into the one tube:

- 6.0 x (N+1) µL of PCR buffer,
- 0.3 x (N+1) µL of TechnoTaq MAX polymerase,

N is a quantity of the samples to be tested taking to account "C-", "C+".

Example: for simultaneous testing of 5 samples, "C-" and "C+" in one PCR run, mark 7 tubes (5 tubes for samples to be tested, 1 tube for "C+" and 1 tube for "C-"). Prepare the mixture of PCR buffer and TechnoTaq MAX polymerase for 8 (7+1) tubes. Mix 48 µL of PCR buffer and 2.4 µL of TechnoTaq MAX polymerase.

8.2.6 Vortex the tube with the mixture of PCR buffer and TechnoTaq MAX polymerase (3-5 seconds) and spin briefly on vortex mixer- (1-3 seconds).



Mixture of PCR buffer and TechnoTaq MAX polymerase must be prepared immediately prior to use.

8.2.7 Add 6.0 µL of PCR buffer and TechnoTaq MAX polymerase mixture into each tube with PCR mix.



Follow the steps listed in pp. 8.2.8 – 8.2.13 within two hours after addition of PCR buffer and TechnoTaq MAX polymerase mixture to PCR mix.

8.2.8 Vortex the tubes with samples, "C+" and "C-" for 3-5 seconds and spin down drops for 1-3 seconds.



In case of using **PREP-GS DNA Extraction Kit**. After vortexing centrifuge the tubes with the DNA preparation at RCF(g) 16000 for one minute to precipitate the sorbent. If, after isolation, the supernatant containing the isolated DNA was transferred to new tubes, centrifugation is carried out for 1-3 seconds in a vortex mixer.



Open the cap of the tube, add DNA sample (or control sample), then close the tube before proceeding to the next tube to prevent contamination. Close the tubes tightly. Use filter tips.

8.2.9 Add 6.0 µL of DNA sample into corresponding tubes. Do not add DNA into the "C+", "C-" tubes.

8.2.10 Add 6.0 µL of negative control (C-) which passed whole DNA extraction procedure into corresponding tube. Add 6.0 µL of positive control sample (C+) into corresponding tube.

8.2.11 Spin tubes for 3-5 seconds (when using the Rotor-Gene Q thermal cycler, centrifugation is not required).

8.2.12 Set the tubes into the Real-time Thermal Cycler.

8.2.13 Launch the operating software for DT instrument³. Add corresponding test⁴, specify the number and ID's of the samples, positive and negative control samples. Specify the position of the tubes in the thermal unit (see 8.2.12) and run PCR. See Tables 9-13.

³ Please, apply to Operation Manual for DTprime and DTlite Real-Time PCR instruments PART II.

⁴ Instructions for uploading "files with test parameters" can be found on "DNA-Technology's" website <https://www.dna-technology.com/assaylibrary>.

Table 10. The PCR program for DTLite and DTprime Thermal Cyclers for package U

Step	Temperature, °C	Min.	Sec.	Number of cycles	Optical measurement	Type of the step
1	80	0	5	15		Cycle
	94	0	5			
2	94	5	00	1		Cycle
3	94	0	30	5		Cycle
	64	0	15		√	
4	94	0	10	45		Cycle
	64	0	15		√	
5	94	0	5	1		Cycle
6	10 ¹	...	...	Holding		Holding
√ - optical measurement						
¹ – holding at 25°C is allowed						

Table 11. The PCR program for iCycler iQ thermal cycler (with persistent well factor) for package U

Cycle	Repeats	Step	Dwell time	Setpoint, °C	PCR/Melt Data Acquisition
1	1				
		1	1 min	80	
		2	5 min	94	
2	5				
		1	30 sec	94	
		2	45 sec	64	
3	45				
		1	10 sec	94	
		2	45 sec	64	Real Time
4	...	...	...	10	Storage

Table 12. The PCR program for iCycler iQ thermal cycler (with dynamic well factor) for package U

Cycle	Repeats	Step	Dwell time	Setpoint, °C	PCR/Melt Data Acquisition
dynamicwf.tmo program					
1	1				
		1	1 min	80	
		2	5 min	94	
2	5				
		1	30 sec	94	
		2	45 sec	64	
3	2				
		1	30 sec	80	Real Time
PCR program					
4	45				
		1	10 sec	94	
		2	45 sec	64	Real Time
5	...	...	...	10	Storage

Table 13. The PCR program for Rotor-Gene Q thermal cycler for package U

Cycling	Temperature	Hold time	Cycle repeats
Cycling	80 deg	60 sec	1 time
	94 deg	300 sec	
Cycling 2	94 deg	30 sec	5 times
	57 deg*	15 sec	
Cycling 3	94 deg	10 sec	45 times
	57 deg*	15 sec	
* Take the measurement			

8.3 Preparing PCR using DTStream (only for package U)



The reagents and tubes should be kept away from direct sun light.

8.3.1 Vortex the tube with PCR mix for 3-5 seconds, then spin for 1-3 seconds to collect the drops.

8.3.2 Vortex the TechnoTaq MAX polymerase and PCR buffer for 3-5 seconds, then spin for 1-3 seconds.



TechnoTaq MAX polymerase should be got out from the freezer immediately prior to use.

8.3.3 Prepare the mixture of PCR buffer and TechnoTaq MAX polymerase. Follow the DTstream instructions.

8.3.4 Vortex the tube with the mixture of PCR buffer and TechnoTaq MAX polymerase (3-5 seconds) and spin briefly (1-3 seconds).

8.3.5 Vortex the tubes with samples and "C-" and "C+" for 3-5 seconds and spin down the drops for 1-3 seconds.



In case of using **PREP-GS DNA Extraction Kit**. After vortexing centrifuge the tubes with the DNA preparation at RCF(g) 16000 for one minute to precipitate the sorbent. If, after isolation, the supernatant containing the isolated DNA was transferred to new tubes, centrifugation is carried out for 1-3 seconds in a vortex mixer.

8.3.6 Set tubes with PCR mix, mixture of PCR buffer and TechnoTaq MAX, analyzed samples, positive and negative controls and PCR microtray to the DTstream and dispense the components according to the instruction manual.

8.3.7 After the end of dosing program on DTstream put the PCR microtray without shaking on the working table of DTpack sealing device.

8.3.8 Run the process of sealing of PCR microtray according to the user manual of DTpack sealing device.

8.3.9 Centrifuge the microtray at RCF(g) 100 for 30 seconds.

8.3.10 Launch the operating software for DT instrument⁵. Add corresponding test⁶, specify the number and ID's of the samples, positive and negative control samples. Specify the position of the tubes in the thermal unit (see 8.3.9) and run PCR. See Tables 9-13.

⁵ Please, apply to Operation Manual for DTprime and DTLite Real-Time PCR instruments PART II.

⁶ Instructions for uploading "files with test parameters" can be found on "DNA-Technology's" website <https://www.dna-technology.com/assaylibrary>.

9. CONTROLS

The **Helicobacter pylori REAL-TIME PCR Detection Kit** contains positive control sample. Positive control is a cloned part of the *Helicobacter pilory* genome. It is produced with genetic engineering techniques and characterized by automatic DNA sequencing. The PCR mix from the kit includes the Internal control (IC). IC is an artificial plasmid intended to assess the quality of PCR performance.



A negative control sample should go through all stages of DNA extraction. Physiological saline solution can be used as a negative control sample in volumes indicated in supplied instructions.

The test result is considered valid when:

- the exponential growth of the fluorescence level for the specific product is present, in this case the internal control is not taken into account;
- the exponential growth of the fluorescence level for the specific product is absent and for internal control is present.

The test result is considered invalid when the exponential growth of the fluorescence level for the specific product and for internal control is not observed.

If positive control (C+) does **not** express growing fluorescence of the specific product or positive result, it is required to repeat the whole test. It may be caused by inhibitors, operation error or violation of storage and handling.

If negative control (C-) expresses growing fluorescence of the specific product or positive result, all tests of the current batch are considered false. Decontamination is required.

10. DATA ANALYSIS

In case of using DNA-Technology made Real-Time PCR Thermal Cyclers, the analysis is performed automatically. In all other cases, the analysis is based on the presence or absence of specific signal.

In the samples containing *Helicobacter pilory* DNA (specific product), the Real-Time PCR Thermal Cycler registers the expressed growing fluorescence of specific product, the amplification result of the internal control is not taken into account.

In the samples free of *Helicobacter pilory* DNA, the Real-Time PCR Thermal Cycler registers the expressed growing fluorescence of the internal control and its absence for the specific product.

When the unseen expressed growing fluorescence or negative result of both in the specific product and the internal control, the result of amplification is considered as uncertain. It may be due to inhibitors, incorrect performance, non-compliance of the amplification temperatures, etc. In this case, amplification, or DNA extraction, or collecting of clinical material are required to be repeated.

In case the result for negative control is defined as positive, the whole experiment should be considered false. The retesting and decontamination are required.

The controls should be also considered to exclude false positive and false negative results (see p. 9 of the current manual).

11. SPECIFICATIONS

a. The analytical specificity of the **Helicobacter pylori REAL-TIME PCR Detection Kit** was assessed by bioinformatics analysis using available on-line databases with up-to-date comprehensive genetic information. The specific oligonucleotides used in the test were checked against GenBank database sequences. None of the sequences showed sufficient similarity for unspecific detection.

The samples with *Helicobacter pilory* DNA are to be registered positive for specific product (a fragment of the *Helicobacter pilory* genome). The samples free of *Helicobacter pilory* DNA are to be registered negative for specific product and positive for internal control.

There are not non-specific positive results of amplification of DNA sample in the presence of: *Campylobacter* genus: *Campylobacter showae*, *Campylobacter hyointestinalis*, *Campylobacter rectus*, *Campylobacter gracilis*, *Campylobacter concisus*, *Campylobacter hominis*, *Campylobacter fetus*, *Campylobacter coli*, *Campylobacter jejuni*, *Campylobacter lari*, *Campylobacter upsaliensis*, *Campylobacter curvus*, *Campylobacter sputorum*, *Helicobacter fennelliae*, *Bacteroides ureolyticus*;

Helicobacter genus: *Helicobacter mustelae*, *Helicobacter muridarum*, *Helicobacter canis*, *Helicobacter bilis*, *Helicobacter cinaedi*, *Helicobacter hepaticus*, *Helicobacter trogonum*, *Helicobacter ganmani*, *Helicobacter winghamensis*, *Helicobacter canadensis*, *Helicobacter pullorum*, *Helicobacter bizzozeronii*, *Helicobacter heilmannii*, *Flexispira rappini*;

Pseudomonas spp. genus: *Pseudomonas putida*, *Pseudomonas aeruginosa*, *Pseudomonas fluorescens*;

Enterobacterales genus: *Cronobacter sakazakii*, *Citrobacter koseri*, *Citrobacter braakii*, *Citrobacter freundii*, *Salmonella enterica*, *Salmonella enteritidis*, *Escherichia coli*, *Escherichia fergusonii*, *Shigella flexneri*, *Shigella dysenteriae*, *Serratia rubidaea*, *Serratia marcescens*, *Enterobacter aerogenes*, *Klebsiella oxytoca*, *Enterobacter cancerogenus*, *Enterobacter cloacae*, *Enterobacter asburiae*, *Hafnia alvei*, *Yersinia pseudotuberculosis*, *Yersinia enterocolitica*, *Providencia alcalifaciens*, *Morganella morganii*, *Proteus mirabilis*, *Proteus vulgaris*, *Aeromonas hydrophila*, *Aeromonas punctate*;

and human DNA in concentration up to 1.0×10^8 copies/mL of sample.

b. The analytical sensitivity of the **Helicobacter pylori REAL-TIME PCR Detection Kit** is 5 copies of *Helicobacter pylori* DNA per amplification tube. Sensitivity is determined by the analysis of serial dilutions of the laboratory control sample (LCS). 94 tests were made for each concentration.

The concentration of LCS, copies per amplification tube	Number of repetitions	Number of positive results	% of positive results
5	94	94	100
2	94	79	84

c. Diagnostic characteristics

Number of samples (n) - 100;

Diagnostic sensitivity (95% CI) – 100% (92.89 – 100%);

Diagnostic specificity (95% CI) – 100% (92.89 – 100%).



The claimed specifications are guaranteed when DNA extraction is performed with **PREP-NA** **REF**P-002/1EU and **PREP-GS** **REF**P-003/1EU extraction kits.

12. TROUBLESHOOTING

Table 14. Troubleshooting

	Result	Possible cause	Solution
C+	-	Operation error PCR inhibition Violation of storage and handling requirements	Repeat whole test Dispose current batch
C-	+	Contamination	Dispose current batch Perform decontamination procedures
IC	Invalid	PCR inhibition	Repeat whole test Resample

If you face to any undescribed issues contact our customer service department regarding quality issues with the kit:

Phone: +7(495)640.16.93

E-mail: hotline@dna-technology.ru

<https://www.dna-technology.com/support>

13. QUALITY CONTROL

The quality control procedures performed in accordance with ISO 9001:2015 and ISO 13485:2016:

- observation of quality management in manufacturing of products;
- creation of values for customers;
- maintenance of the best service quality and customer management.

Contact our customer service by quality issues of **Helicobacter pylori REAL-TIME PCR Detection Kit**.

Technical support:

E-mail: hotline@dna-technology.ru

<https://www.dna-technology.com>

Manufacturer: "DNA-Technology Research & Production", LLC,

142281, Russia, Moscow Region,

Protvino, Zheleznodorozhnaya Street, 20

Phone/fax: +7(495) 640.17.71

E-mail: info@dna-technology.com

<https://www.dna-technology.com>

Seller: "DNA-Technology" LLC,

117587, Russia, Moscow,

int. ter. Municipal District Chertanovo Severnoye,

Varshavskoye shosse, 125 Zh, building 5, floor 1, office 12;

Phone/fax: +7(495) 640.17.71

E-mail: info@dna-technology.com

<https://www.dna-technology.com>

14. KEY TO SYMBOLS

	<i>In vitro</i> diagnostic medical device		Date of manufacture
	Temperature limit		Consult instructions for use
	Contains sufficient for <n> tests		Catalogue number
	Use-by date		Manufacturer
	Batch code		Keep away from sunlight
	Caution		Version
	Do not reuse		Positive control
	Non-sterile		

REF

R1-P501-S3/9EU
R1-P501-23/9EU
R1-P501-UA/9EU

VER

1534.2026.06.05