


For professional use only

STOR-M transport medium

INSTRUCTION FOR USE



OBELIS S.A

Registered Address:

Bd. Général Wahis, 53

1030 Brussels, Belgium

Tel: +32.2.732.59.54

Fax: +32.2.732.60.03

E-mail: mail@obelis.net

<http://www.obelis.net>



"DNA-Technology Research &
Production", LLC,

142281, Russia,

Moscow Region, Protvino,

Zheleznodorozhnaya Street, 20

Phone/fax: +7(495) 640.17.71

E-mail: info@dna-technology.com

<https://www.dna-technology.com>

Customer service department

E-mail: hotline@dna-technology.ru

TABLE OF CONTENTS

1. INTENDED USE	3
2. METHOD	3
3. CONTENT	3
4. REAGENTS AND EQUIPMENT REQUIRED BUT NOT PROVIDED	3
5. TRANSPORT AND STORAGE CONDITIONS	4
6. WARNINGS AND PRECAUTIONS	4
7. SAMPLES	5
8. PROCEDURE	6
9. SPECIFICATIONS	7
10. QUALITY CONTROL	8
11. KEY TO SYMBOLS	9

1. INTENDED USE

The **STOR-M transport medium** is intended for transport and storage of human biological samples (scrapes/swabs of epithelial cells from urogenital tract, oropharynx, nasopharynx, rectum, skin, conjunctiva of the eye), including those containing an impurity of mucus, followed by nucleic acids analysis (human and microbial DNA, viral RNA) by polymerase chain reaction method.

This medical device is an auxiliary agent in clinical laboratory diagnostics.

The application of the kit does not depend on population and demographic aspects. There are no contradictions for use of the **STOR-M transport medium**.

The **STOR-M transport medium** can be used in clinical and diagnostic laboratories of medical institutions and research practice.

Potential users: personnel qualified in molecular diagnostics methods and working in the clinical and diagnostic laboratory.

It is necessary to apply the kit only as directed in this instruction for use.

2. METHOD

The **STOR-M transport medium** is a ready-to-use water-salt transparent, colorless solution with the addition of a preservative, mucolytic and flavoring agent. The preservative prevents the growth of non-specific microorganisms, and mucolytic acts on disulfide bonds of mucopolysaccharides, thereby diluting mucus.

3. CONTENT

The detailed description of content is represented in Table 1.

Table 1. The **STOR-M transport medium** content, package S (standard) for P-910-1/1EU

Reagent	Description	Total volume	Amount
Transport medium	Transparent colorless liquid	50 mL (500 µL in each tube)	100 tubes

All components are ready to use and do not require additional preparation for operation.

The kit is intended for single use and designed for 100 tests for **STOR-M transport medium**.

4. REAGENTS AND EQUIPMENT REQUIRED BUT NOT PROVIDED

4.1. Specimen collection

- Specimen collection swabs: sterile single use swabs, cytobrushes, cotton swabs e.t.c for sampling of biomaterial.

4.2. Preparation for NA extraction

- Biological safety cabinet class II;
- Vortex mixer;
- Refrigerator;
- NA extraction kit;
- High speed centrifuge (RCF(g) no less than 12000);
- Tube rack for 1.5 mL tubes;
- Container for used pipette tips, tubes and other consumables;
- Single channel pipettes (dispensers covering 100-1000 µL volume range);

- RNase and DNase free filtered pipette tips (volume 1000 µL);
- Powder-free surgical gloves;
- Disinfectant solution.

5. TRANSPORT AND STORAGE CONDITIONS

Expiry date – 12 months from the date of production.

All components of the **STOR-M transport medium** must be stored at temperatures from 2 °C to 8 °C over the storage period. The excessive temperature can be detrimental to product performance.

Transportation of the kit is allowed in thermal containers with icepacks by all types of covered transport at temperatures from 2 °C to 25 °C inside the container, but for no longer than 14 days.

Shelf-life of the kit following the first opening of the primary container: components of the kit should be stored at temperatures from 2 °C to 8 °C during the storage period.

The kit stored in under undue regime should not be used.

An expired the **STOR-M transport medium** should not be used.

We strongly recommend to follow the given instructions in order to obtain accurate and reliable results.

The conformity of the **STOR-M transport medium** to the prescribed technical requirements is subject to compliance of storage, transportation and handling conditions recommended by manufacturer.

Contact our official representative in EU by quality issues of the **STOR-M transport medium**.

6. WARNINGS AND PRECAUTIONS

Only personnel trained in the methods of molecular diagnostics and the rules of work in the clinical and diagnostic laboratory are allowed to work with the kit.

Handle and dispose all biological samples, reagents and materials used to carry out the assay as if they were able to transmit infective agents. The samples must be exclusively employed for certain type of analysis. Samples must be handled under a laminar flow hood. Tubes containing different samples must never be opened at the same time. Pipettes used to handle samples must be exclusively employed for this specific purpose. The pipettes must be of the positive dispensation type or be used with aerosol filter tips. The tips employed must be sterile, free from the DNases and RNases, free from DNA and RNA. The reagents must be handled under a laminar flow hood. The reagents required for amplification must be prepared in such a way that they can be used in a single session. Pipettes used to handle reagents must be exclusively employed for this specific purpose. The pipettes must be of the positive dispensation type or be used with aerosol filter tips. The tips employed must be sterile, free from the DNases and RNases, free from DNA and RNA. Avoid direct contact with the biological samples reagents and materials used to carry out the assay. Wear powder-free surgical gloves. Wear protective clothing (work clothes and personal protective equipment) working with microorganisms classified as particularly pathogenic. The protective clothing and personal protective equipment must comply with the work to be performed and health and safety requirements. Avoid producing spills or aerosol. Any material being exposed to biological samples must be treated for at least 30 minutes with disinfecting solution or autoclaved for 1 hour at 121 °C before disposal.

Molecular biology procedures, such as nucleic acids extraction, reverse transcription, PCR-amplification and detection require qualified staff to avoid the risk of erroneous results, especially due to the degradation of nucleic acids contained in the samples or sample contamination by amplification products.

All the liquid solutions are designed for single use and can not be used more than once. Plastic tubes do not contain phthalates. Do not breathe gas/fumes/vapor/spray produced by the components of the kit. Do not eat/drink components of the kit. Avoid contact with eyes. Only use the reagents provided in the kit and those recommended by manufacturer. Do not mix reagents from different batches. Do not use reagents from third party manufacturers' kits. All laboratory equipment, including dispensers, test tube

racks, laboratory glassware, lab coats, bouffant caps, etc., as well as reagents should be strictly stationary. It is not allowed to move them from one room to another. Waste materials are disposed of in accordance with local and national standards. All surfaces in the laboratory (work tables, test tube racks, equipment, etc.) must be treated daily with disinfecting solution.

Emergency actions

Eye Contact: If any component of this kit enters the eyes, wash eyes gently under potable running water for 15 minutes or longer, making sure that the eyelids are held open. If pain or irritation occurs, obtain medical attention.

Skin Contact: If any component of this kit contacts the skin and causes discomfort, remove any contaminated clothing. Wash affected area with plenty of soap and water. If pain or irritation occurs, obtain medical attention.

Ingestion: If any component of this kit is ingested, wash mouth out with water. If irritation or discomfort occurs, obtain medical attention.

Do not use the kit:

- When the transportation and storage conditions are breached;
- When the reagents' appearance does not respond to the kit passport;
- When the kit components packaging is breached;
- After the expiry date provided.

Significant health effects are **NOT** anticipated from routine use of this kit when adhering to the instructions listed in the current manual.

7. SAMPLES

The **STOR-M transport medium** is designed to transport and storage the scrapes/swabs of epithelial cells from urogenital tract, oropharynx, nasopharynx, rectum, skin, conjunctiva of the eye.

General requirements

Remove free separable mucus with a sterile cotton prior to sampling. In case of sampling from several locations, repeat the procedure several times, each time taking a new swab into new different tube.

To prevent contamination, open the tube, add sample, then close the tube before proceeding to the next sample.

Scrapes/swabs of epithelial cells sampling

Order of taking:

1. Open the 1.5 mL tube with **STOR-M transport medium**.
2. Scrape epithelial cells from the corresponding biotope (i.e. urogenital tract, oropharynx, nasopharynx, rectum, skin, conjunctiva of the eye) with a sterile swab.
3. Put the swab into the tube with transport medium and rinse it thoroughly. Avoid spraying of solution.
4. Remove swab from solution, press it to the wall of tube and squeeze the rest of the liquid. Throw out the swab.
5. Close the tube tightly and mark it.

Transportation and storage of the samples

Transport and store samples in **STOR-M transport medium** for subsequent analysis of human and microbial DNA at temperatures from 2 °C to 8 °C for no longer than 3 months.

Transport and store samples in **STOR-M transport medium** for subsequent analysis of viral RNA at temperatures from 2 °C to 8 °C for no longer than 28 days.

Transport and store samples in **STOR-M transport medium** for subsequent analysis of human and microbial DNA at temperatures from 18 °C to 25 °C is acceptable for no longer than 28 days.

Transport and store samples in **STOR-M transport medium** for subsequent analysis of viral RNA at temperatures from 18 °C to 25 °C is acceptable for no longer than 7 days.

8. PROCEDURE

8.1 General recommendations

1. Use only disposable RNase and DNase free filtered pipette tips.
2. When adding the solution into sample, do not touch the walls of the tubes with the tip. If the tip has touched the wall of the tube, change the tip. Tip should be changed each time when you take out solution from the tube with sample.
3. To prevent contamination, open the tube, add sample/reagent, then close the tube before proceeding to the next sample/reagent.

8.2 Preparation for NA extraction

- 8.2.1 Centrifuge samples in transport medium at RCF(g) 12000 for 10 minutes.
- 8.2.2 Remove the supernatant, leaving the approximately 100 µL (precipitate+liquid fraction) in the tube.



If for sample pretreatment centrifugation is not needed, p.8.2.1 and 8.2.2 are not performed.



For scrape/swab from oropharynx and nasopharynx for subsequent analysis of viral RNA preliminary centrifugation is not required.

- 8.2.3 Add the lysis buffer in the volume recommended by the NA extraction kit used into the tube with precipitate.
- 8.2.4 Perform NA extraction according to NA extraction kit user manual.
- 8.2.5 To make a negative control sample perform p. 8.2.1 and 8.2.2 for tube with **STOR-M transport medium** which does not contain sample. Or use necessary volume of saline as sample according to NA extraction kit manual.

STOR-M transport medium is compatible with any NA extraction methods.

The recommended kits for nucleic acids extraction:

- **PREP-RAPID DNA Extraction Kit** ("DNA-Technology Research & Production", LLC, Russia);
- **PREP-GS and PREP-GS PLUS DNA Extraction Kits** ("DNA-Technology Research & Production", LLC, Russia);
- **PREP-NA and PREP-NA PLUS DNA/RNA Extraction Kits** ("DNA-Technology Research & Production", LLC, Russia);
- **PREP-MB RAPID DNA Extraction Kit** ("DNA-Technology Research & Production", LLC, Russia);
- **PREP-NA-S DNA/RNA Extraction Kit** ("DNA-Technology Research & Production", LLC, Russia).

9. SPECIFICATIONS

The preservation of DNA in biomaterial samples at different storage conditions

Biomaterial	Preservation of DNA (%)	
	28 days under temperatures from 18 °C to 25 °C	3 months under temperatures from 2 °C to 8 °C
scrapes/swabs of epithelial cells from urogenital tract	76.7	67.5
scrapes/swabs of epithelial cells from oropharynx	79.6	68.6
scrapes/swabs of epithelial cells from nasopharynx	64.9	67.6
scrapes/swabs of epithelial cells from conjunctiva of the eye	64.3	51.6
scrapes/swabs of epithelial cells from rectum	86.4	93.1
scrapes/swabs of epithelial cells from skin	83.6	67.1
Total (average, 95% CI)	75.9 (67.8-85.2) n=50	69.2 (57.8-79.4) n=50

The preservation of viral RNA in biomaterial samples at different storage conditions

Biomaterial	Preservation of RNA (%)	
	7 days under temperatures from 18 °C to 25 °C	28 days under temperatures from 2 °C to 8 °C
scrapes/swabs of epithelial cells from oropharynx (SARS-CoV-2)	No losses. The sample is applicable for RT-PCR	No losses. The sample is applicable for RT-PCR
scrapes/swabs of epithelial cells from nasopharynx (HRV)	No losses. The sample is applicable for RT-PCR	No losses. The sample is applicable for RT-PCR
Total	n=7	n=9

10. QUALITY CONTROL

“DNA-Technology Research&Production”, LLC declares that the above mentioned products meet the provision of the Council Directive 98/79/EC for *in vitro* Diagnostic Medical Devices. The quality control procedures performed in accordance with ISO 9001:2015 and ISO 13485:2016:

- observation of quality management in manufacturing of IVDD products;
- creation of values for customers;
- maintenance of the best service quality and customer management.

Contact our official representative in EU by quality issues of **STOR-M transport medium**.

Technical support:

E-mail: hotline@dna-technology.ru

<https://www.dna-technology.com>

Manufacturer: "DNA-Technology Research & Production", LLC,

142281, Russia, Moscow Region,

Protvino, Zheleznodorozhnaya Street, 20

Phone/fax: +7(495) 640.17.71

E-mail: info@dna-technology.com

<https://www.dna-technology.com>

Seller: “DNA-Technology” LLC,

117587, Russia, Moscow,

int. ter. Municipal District Chertanovo Severnoye,

Varshavskoye shosse, 125 Zh, building 5, floor 1, office 12;

Phone/fax: +7(495) 640.17.71

E-mail: info@dna-technology.com

<https://www.dna-technology.com>

Authorized representative in EU:

OBELIS S.A

Registered Address:

Bd. Général Wahis, 53

1030 Brussels, Belgium














Tel: +32.2.732.59.54

Fax: +32.2.732.60.03

E-mail: mail@obelis.net

<http://www.obelis.net>

11. KEY TO SYMBOLS

	<i>In vitro</i> diagnostic medical device		Date of manufacture
	Temperature limit		Consult instructions for use
	Contains sufficient for <n> tests		Catalogue number
	Use-by date		Manufacturer
	Batch code		Version
	Caution		Non-sterile
	Authorized representative in the European Community		

REF

P-910-1/1EU

VER

600-2.2022.08.30