

1130 2025-06-23

**For research use only**

## **Streptococcus pneumoniae REAL-TIME PCR Detection Kit**

### **INSTRUCTION FOR USE**



"DNA-Technology Research &  
Production", LLC,

142281, Russia,

Moscow Region, Protvino,

Zheleznodorozhnaya Street, 20

Phone/fax: +7(495) 640.17.71

E-mail: [info@dna-technology.com](mailto:info@dna-technology.com)

<https://www.dna-technology.com>

Customer service department

E-mail: [hotline@dna-technology.ru](mailto:hotline@dna-technology.ru)



R1-P412-S3/4ER  
R1-P412-23/4ER  
R1-P412-UA/9ER



1130.2025.06.23

## TABLE OF CONTENTS

1.	INTENDED USE.....	3
2.	METHOD .....	3
3.	CONTENT .....	4
4.	REAGENTS AND EQUIPMENT REQUIRED BUT NOT PROVIDED .....	5
5.	TRANSPORT AND STORAGE CONDITIONS .....	7
6.	WARNINGS AND PRECAUTIONS .....	7
7.	SAMPLES .....	9
8.	PROCEDURE .....	12
9.	CONTROLS.....	18
10.	DATA ANALYSIS .....	19
11.	SPECIFICATIONS.....	20
12.	TROUBLESHOOTING .....	21
13.	QUALITY CONTROL .....	22
14.	KEY TO SYMBOLS.....	23
	Annex A .....	24

## 1. INTENDED USE

The ***Streptococcus pneumoniae* REAL-TIME PCR Detection Kit** is an *in vitro* Nucleic Acid Test (NAT) – pathogen-detection-based product. The ***Streptococcus pneumoniae* REAL-TIME PCR Detection Kit** is designed to detect DNA of *Streptococcus pneumoniae* in human biological material (nasopharyngeal, oropharyngeal smears; bronchoalveolar lavage; endotracheal, nasopharyngeal aspirate; phlegm; pleural fluid) by real-time PCR.

The application of the kit does not depend on population and demographic aspects. There are no contradictions for use of the ***Streptococcus pneumoniae* REAL-TIME PCR Detection Kit**.

The ***Streptococcus pneumoniae* REAL-TIME PCR Detection Kit** can be used in research practice.

Potential users: qualified personnel trained in research methods and rules of work in the laboratory.

It is necessary to apply the kit only as directed in this instruction for use.

## 2. METHOD

**Method:** polymerase chain reaction (PCR) with real time results detection; qualitative multiplex analysis.

The implemented PCR method is based on amplification of a target DNA sequence. The amplification process consists of a series of repeated cycles of temperature denaturation of DNA, annealing of primers with complementary sequences, and subsequent completion of the polynucleotide chains from these primers with Taq polymerase.

To increase sensitivity and specificity of the amplification reaction, the use of a “hot” start is provided. For package S, “hot” start is provided by reaction mixture preparation consisting of two layers separated by a layer of paraffin. The polymerase chain reaction starts only when paraffin is melted. This eliminates the nonspecific anchoring of primers with the DNA target at lower temperatures. “Hot” start for package U is provided by using polymerase which activity is blocked by antibodies. The activation of the enzyme occurs only after preheating the reaction mixture at 94 °C. This eliminates the nonspecific anchoring of primers with the DNA target at lower temperatures.

DNA probes each containing a fluorescent label and a fluorescence quencher are introduced into the PCR-mix. When a specific product is formed, the DNA probe is destroyed and the quencher stops affecting the fluorescent label, which leads to an increase in the fluorescence level. The number of destroyed probes (and hence the fluorescence level) increases in proportion to the number of specific amplification products formed. The fluorescence level is measured at each amplification cycle in real time.

PCR-mix includes internal control (IC) designed to control the quality of polymerase chain reaction.

The DNA probe used to detect the *Streptococcus pneumoniae* amplification product includes fluorescent dye Fam (Green). The DNA probe used to detect the internal control amplification product includes fluorescent dye Hex (Yellow, Vic).

Table 1 shows the detection channels of amplification products.

Table 1. Detection channels of amplification products

Fam/Green	Hex/Yellow/Vic	Rox/Orange	Cy5/Red	Cy5.5/Crimson
<i>Streptococcus pneumoniae</i>	IC	-	-	-

The automatic analysis available on “DNA-Technology” made instruments: DTlite, DTprime or DTprime II real-time thermal cyclers for **Streptococcus pneumoniae REAL-TIME PCR Detection Kit** (see the catalogue at <https://www.dna-technology.com> to see available supply options). The current version of the software is available for download at <https://www.dna-technology.com/software>.

The **Streptococcus pneumoniae REAL-TIME PCR Detection Kit** is also approved for use with Rotor-Gene Q (Qiagen), CFX96 (Bio-Rad) and Applied Biosystems QuantStudio 5 (Life Technologies Holdings Pte. Ltd) real-time thermal cyclers.

### 3. CONTENT

The **Streptococcus pneumoniae REAL-TIME PCR Detection Kit** contents is represented in Tables 2 - 4.

Table 2. The **Streptococcus pneumoniae REAL-TIME PCR Detection Kit** content, package S, strips, for R1-P412-S3/4ER

Reagent	Description	Total volume	Amount
Paraffin sealed PCR-mix	Colorless or pink transparent liquid under waxy white fraction	960 µL (20 µL in each)	tubes, 6 strips of 8
Taq-polymerase solution	Colorless transparent liquid	500 µL	1 tube
Mineral oil	Colorless transparent viscous oily liquid	1.0 mL	1 tube
Positive control <sup>1</sup>	Colorless transparent liquid	130 µL	1 tube
Strip caps	6 strips of 8		

Table 3. The **Streptococcus pneumoniae REAL-TIME PCR Detection Kit** content, package S, tubes, for R1-P412-23/4ER

Reagent	Description	Total volume	Amount
Paraffin sealed PCR-mix	Colorless or pink transparent liquid under waxy white fraction	960 µL (20 µL in each)	48 individual tubes
Taq-polymerase solution	Colorless transparent liquid	500 µL	1 tube
Mineral oil	Colorless transparent viscous oily liquid	1.0 mL	1 tube
Positive control <sup>1</sup>	Colorless transparent liquid	130 µL	1 tube

Table 4. The **Streptococcus pneumoniae REAL-TIME PCR Detection Kit** content, package U, for R1-P412-UA/9ER

Reagent	Description	Total volume	Amount
PCR-mix	Colorless or pink transparent liquid	600 µL	1 tube
TechnoTaq MAX polymerase	Colorless transparent viscous liquid	30 µL	1 tube
PCR-buffer	Colorless transparent liquid	600 µL	1 tube
Positive control <sup>1</sup>	Colorless transparent liquid	130 µL	1 tube

All components are ready to use and do not require additional preparation for operation.

<sup>1</sup> - marking as C+ is allowed

The ***Streptococcus pneumoniae* REAL-TIME PCR Detection Kit** (package S) is designed for 48 tests (no more than 12 runs), including analysis of test samples, negative controls and positive controls.

The ***Streptococcus pneumoniae* REAL-TIME PCR Detection Kit** (package U) is designed for 96 tests with at least 5 samples per run (3 test samples, negative control and positive control).

#### 4. REAGENTS AND EQUIPMENT REQUIRED BUT NOT PROVIDED

The following equipment, reagents and consumables are required:

Equipment, reagents and consumables	Package S		Package U, dosing	
	strips <sup>1</sup>	tubes	manual	automated
UV PCR cabinet	yes	yes	yes	yes
Real-time detecting thermal cycler <sup>2</sup>	yes	yes	yes	yes <sup>3</sup>
Vortex mixer	yes	yes	yes	yes
Vortex rotor for 0.2 mL strips	yes	no	no	no
Refrigerator with freezer	yes	yes	yes	yes
Tube rack for 1.5 mL tubes	yes	yes	yes	yes
Tube rack for 0.2 mL tubes	no	yes	yes <sup>4</sup>	no
Tube rack for 0.2 mL tubes or strips	yes	no	no	no
Single channel pipettes (dispensers covering 2.0-1000 µL volume range)	yes	yes	yes	yes
RNase and DNase free filtered pipette tips (volume 20 µL, 200 µL, 1000 µL)	yes	yes	yes	yes
Pipette rack	yes	yes	yes	yes
RNase and DNase free 1.5 mL microfuge tubes with caps	yes	yes	yes	yes
RNase and DNase free 0.2 mL PCR tubes or 96-well PCR microplate <sup>5</sup>	no	no	yes	no
Powder-free surgical gloves	yes	yes	yes	yes
Container for used pipette tips, tubes and other consumables	yes	yes	yes	yes
DTstream dosing instrument, version 12M1 or 15M1	no	no	no	yes
RNase and DNase free filtered pipette tips (volume 200 µL) for DTstream	no	no	no	yes
DTpack plate sealing device	no	no	yes <sup>6</sup>	yes
Centrifuge for microplates (RCF(g) at least 100)	no	no	yes <sup>6</sup>	yes
Polymer thermal film for microplate sealing	no	no	yes <sup>6</sup>	yes
384-well PCR microplate	no	no	no	yes
Physiological saline solution 0.9% NaCl (sterile) (if necessary)				
Transport medium (if necessary), the following are recommended: - <b>STOR-F</b> transport medium for biomaterial samples.				

Equipment, reagents and consumables	Package S		Package U, dosing	
	strips <sup>1</sup>	tubes	manual	automated
NA extraction reagent kits <sup>7</sup> , the following are recommended: <ul style="list-style-type: none"> <li>– <b>PREP-NA, PREP-NA PLUS</b> <span style="border: 1px solid black; padding: 0 2px;">REF</span> P-034-N/1EU, P-036-N/1EU, P-036-N/2EU;</li> <li>– <b>PREP-RAPID</b> <span style="border: 1px solid black; padding: 0 2px;">REF</span> P-001/1EU;</li> <li>– <b>PREP-MB RAPID</b> <span style="border: 1px solid black; padding: 0 2px;">REF</span> P-116-A/8EU;</li> <li>– <b>PREP-OPTIMA</b> <span style="border: 1px solid black; padding: 0 2px;">REF</span> P-016-N/2EU, P-016-1/2EU, P-015-N/2EU;</li> <li>– <b>PREP-MB-RAPID II</b> <span style="border: 1px solid black; padding: 0 2px;">REF</span> P-122-A/9EU, P-122-N/9EU, P-122-P/9EU, P-124-P/9EU.</li> </ul>				
<b>Additional equipment for pretreatment of bronchoalveolar lavage; pleural fluid; nasopharyngeal and endotracheal aspirate:</b>				
Biological safety cabinet class II				
RNase and DNase free 1.5 mL microfuge tubes with caps				
Centrifuge for 1.5 mL tubes, RCF(g) at least 12,000				
Disinfectant solution				
<b>Notes:</b> <ul style="list-style-type: none"> <li><sup>1</sup> – not used for Rotor-Gene Q thermal cyclers</li> <li><sup>2</sup> – hereinafter – detecting thermal cyclers; the required parameters are indicated below</li> <li><sup>3</sup> – DTprime *X*, DTprime II *X* are validated</li> <li><sup>4</sup> – only if using tubes</li> <li><sup>5</sup> – not used for DTlite and Rotor-Gene Q thermal cyclers</li> <li><sup>6</sup> – only if using microplates</li> <li><sup>7</sup> – choice of extraction kit is determined by biomaterial type</li> </ul>				

The following detecting thermal cyclers are validated for work with the **Streptococcus pneumoniae REAL-TIME PCR Detection Kit**:

- DTprime in DTprime \*M\* modification (manufactured by “DNA-Technology R&P”, LLC), hereinafter – DTprime;
- DTprime II in DTprime II \*M\* modification (manufactured by “DNA-Technology R&P”, LLC), hereinafter – DTprime II;
- DTprime in DTprime \*X\* modification (manufactured by “DNA-Technology R&P”, LLC), hereinafter – DTprime \*X\* (only for package U, automated dosing);
- DTprime II in DTprime II \*X\* modification (manufactured by “DNA-Technology R&P”, LLC), hereinafter – DTprime II \*X\* (only for package U, automated dosing);
- DTlite in DTlite \*S\* modification (manufactured by “DNA-Technology R&P”, LLC), hereinafter – DTlite (only for package S and package U, manual dosing, tubes);
- Rotor-Gene Q, QIAGEN GmbH, Germany (only for package S, tubes and package U, manual dosing, tubes), hereinafter – Rotor-Gene Q;
- CFX96 (Optical Reaction Module CFX96) (manufactured by Bio-Rad Laboratories, USA), hereinafter – CFX96;
- Applied Biosystems QuantStudio 5 (Life Technologies Holdings Ltd., Singapore), hereinafter – Applied Biosystems QuantStudio 5.

For the use of detecting thermal cyclers other than those listed above, please consult the reagent kit manufacturer for consultation.

Software:

The most recent version of the DT thermal cyclers software can be downloaded from <https://www.dna-technology.com/software>.

The OS supported: all versions of Windows starting from 7.

## **5. TRANSPORT AND STORAGE CONDITIONS**

Expiry date – 12 months from the date of production.

### **5.1. Storage conditions**

#### **5.1.1. Package S**

- All components of the reagent kit must be stored in a refrigerator or a cooling chamber at the temperature from 2 °C to 8 °C throughout the shelf life of the kit.
- Paraffin sealed PCR-mix must be kept away from light.

#### **5.1.2. Package U**

- All components of the reagent kit, except for TechnoTaq MAX polymerase, must be stored in a refrigerator or a cooling chamber at the temperature from 2 °C to 8 °C throughout the shelf life of the kit.
- PCR-mix must be kept away from light.
- TechnoTaq MAX polymerase must be stored in a freezer at the temperature from minus 22 °C to minus 18 °C throughout the shelf life of the kit.

**WARNING!** The excessive temperature and light can be detrimental to product performance.

### **5.2. Transport conditions**

Transportation of the reagent kit is carried out in thermoboxes with ice packs by all types of roofed transport at the temperature inside the container corresponding to the storage conditions of the kit components.

#### **5.2.1. Package S**

- It is allowed to transport the kit in thermoboxes with ice packs by all types of roofed transport at the temperature inside the container from 2 °C to 25 °C for no longer than 5 days.

#### **5.2.2. Package U**

- It is allowed to transport the kit, except for TechnoTaq MAX polymerase, in thermoboxes with ice packs by all types of roofed transport at the temperature inside the container from 2 °C to 25 °C for no longer than 5 days.
- It is allowed to transport the TechnoTaq MAX polymerase in thermoboxes with ice packs by all types of roofed transport at the temperature inside the container up to 25 °C for no longer than 5 days.

**WARNING!** Reagent kits transported with violation of temperature conditions must not be used.

### **5.3. Shelf-life of the kit following the first opening of the primary container**

#### **5.3.1. Package S**

- All components of the kit must be stored in a refrigerator or a cooling chamber at temperatures from 2 °C to 8 °C over the storage period.
- Paraffin sealed PCR-mix must be stored at temperatures from 2 °C to 8 °C and out of light over the storage period.

#### **5.3.2. Package U**

- All components of the kit, except for TechnoTaq MAX polymerase, must be stored in a refrigerator or a cooling chamber at temperatures from 2 °C to 8 °C over the storage period.

- PCR-mix must be stored at temperatures from 2 °C to 8 °C and out of light over the storage period.
- TechnoTaq MAX polymerase must be stored in a freezer at temperatures from minus 22 °C to minus 18 °C over the storage period.

**WARNING!** The kits stored under undue regime must not be used.

An expired kit must not be used.

We strongly recommend to follow the given instructions in order to obtain accurate and reliable results.

The conformity of the **Streptococcus pneumoniae EAL-TIME PCR Detection Kit** to the prescribed technical requirements is subject to compliance of storage, transportation and handling conditions recommended by manufacturer.

## 6. WARNINGS AND PRECAUTIONS

Handle and dispose all biological samples, reagents and materials used to carry out the analysis as if they were able to transmit infective agents. The samples must be exclusively employed for certain type of analysis. Samples must be handled under a laminar flow hood. Tubes containing different samples must never be opened at the same time. Pipettes used to handle samples must be exclusively employed for this specific purpose. The pipettes must be of the positive dispensation type or be used with aerosol filter tips. The tips employed must be sterile, free from the DNases and RNases, free from DNA and RNA. The reagents must be handled under a laminar flow hood. The reagents required for amplification must be prepared in such a way that they can be used in a single session. Pipettes used to handle reagents must be exclusively employed for this specific purpose. The pipettes must be of the positive dispensation type or be used with aerosol filter tips. The tips employed must be sterile, free from the DNases and RNases, free from DNA and RNA. Avoid direct contact with the biological samples reagents and materials used to carry out the analysis. Wear powder-free surgical gloves. Wear protective clothing (work clothes and personal protective equipment) working with microorganisms classified as particularly pathogenic. The protective clothing and personal protective equipment must comply with the work to be performed and health and safety requirements. Avoid producing spills or aerosol. Any material being exposed to biological samples must be treated for at least 30 minutes with disinfecting solution or autoclaved for 1 hour at 121 °C before disposal.

Molecular biology procedures, such as nucleic acids extraction, PCR-amplification and detection require qualified staff to avoid the risk of erroneous results, especially due to the degradation of nucleic acids contained in the samples or sample contamination by amplification products.

All oligonucleotide components are produced by artificial synthesis technology according to internal quality control protocol and do not contain blood or products of blood processing.

Positive control is produced by artificial DNA synthesis technology. Positive control does not include parts of infectious agents.

All the liquid solutions are designed for single use and can not be used more than once in amplification reactions. Plastic tubes do not contain phthalates. Do not breathe gas/fumes/vapor/spray produced by the components of the kit. Do not eat/drink components of the kit. Avoid contact with eyes. Only use the reagents provided in the kit and those recommended by manufacturer. Do not mix reagents from different batches. Do not use reagents from third party manufacturers' kits. All laboratory equipment, including pipettes, test tube racks, laboratory glassware, lab coats, bouffant caps, etc., as well as reagents should be strictly stationary. It is not allowed to move them from one room to another. Equip separate areas for the extraction/preparation of amplification reactions and for the amplification/detection of amplification products. Never introduce an amplification product in the area designed for extraction/preparation of amplification reactions. Wear lab coats, gloves and tools, which are exclusively employed for the extraction/preparation of the amplification reaction and for the amplification/detection of the amplification products. Never transfer lab coats, gloves and tools from the area designed for amplification/detection of the amplification products to the area designed for extraction/preparation of amplification reactions. Amplification products must be handled in such a way as to reduce dispersion into the environment as much as possible, in order to avoid the possibility of contamination. Pipettes used to



handle amplification products must be exclusively employed for this specific purpose. Remove PCR waste only in a closed form. Remove waste materials (tubes, tips) only in a special closed container containing a disinfectant solution. Work surfaces, as well as rooms where NA extraction and PCR are performed, must be irradiated with bactericidal irradiators for 30 minutes before and after the work.

Do not open the tubes after amplification. Waste materials are disposed of in accordance with local and national standards. All surfaces in the laboratory (work tables, test tube racks, equipment, etc.) must be treated daily with disinfecting solution.

### **Emergency actions**

**Inhalation:** Inhalation of the PCR-mix contained within this kit is unlikely, however care should be taken.

**Eye Contact:** If any component of this kit enters the eyes, wash eyes gently under potable running water for 15 minutes or longer, making sure that the eyelids are held open. If pain or irritation occurs, obtain medical attention.

**Skin Contact:** If any component of this kit contacts the skin and causes discomfort, remove any contaminated clothing. Wash affected area with plenty of soap and water. If pain or irritation occurs, obtain medical attention.

**Ingestion:** If any component of this kit is ingested, wash mouth out with water. If irritation or discomfort occurs, obtain medical attention.

Do not use the kit:

- When transportation and storage conditions are breached;
- When the reagents' appearance does not respond to the kit passport;
- When kit components packaging is breached;
- After the expiry date provided.

Significant health effects are **NOT** anticipated from routine use of this kit when adhering to the instructions listed in the current manual.

## **7. SAMPLES**

The ***Streptococcus pneumoniae* REAL-TIME PCR Detection Kit** is designed to detect DNA extracted from nasopharyngeal, oropharyngeal smears; bronchoalveolar lavage; endotracheal, nasopharyngeal aspirate; phlegm; pleural fluid.

### **7.1. General requirements**

- PCR analysis refers to direct methods of laboratory research; therefore, the collection of biological material must be carried out from the site of infection localization. The decision on studying the localization site shall be taken by the physician according to the collected anamnesis and clinical picture of the disease.
- The quality of sampling, sample storage, transport and pretreatment are of great importance for obtaining correct results. Incorrect sampling may lead to unreliable results and, therefore, to the necessity for repeated sampling.
- If biomaterial from several biotopes is required, repeat the procedure using new swab each time you collect biomaterial.
- Use RNase and DNase free filtered tips during biomaterial preparation and NA extraction.
- To prevent contamination, only open the cap of the tube you are working with and close it before proceeding to the next tube.

### **7.2. Interfering substances**

The presence of PCR inhibitors in a sample may cause controversial (uncertain) results. The sign of PCR inhibition is the simultaneous absence of internal control and specific product of amplification.

PCR inhibitors are the presence of hemoglobin in the DNA sample as a result of incomplete removal during the extraction of DNA from a biomaterial sample containing an impurity of blood, as well as the presence of isopropyl alcohol and methyl acetate in the DNA sample as a result of incomplete removal of washing solutions during sample preparation.

The maximum concentration of interfering substances, which do not affect the amplification of the laboratory control sample and internal control: hemoglobin – 0.35 mg/mL DNA sample, isopropyl alcohol – 100 µL/mL DNA sample, methyl acetate – 100 µL/mL DNA sample.

In order to assess possible drug interference, those that could potentially be present in residues in human biological samples taken from the respective biotopes (chlorhexidine bigluconate, Lasolvan® RINO, RINOFLUIMUCIL®, Tizin® Classic, Tantum® Verde Spray, Hexoral® Solution, Berodual®, Salbutamol-Teva, Grippferon® Nasal Drops) were selected.

For all investigated drugs it was shown that there was no effect at concentrations up to 10% in the biomaterial sample.

### **7.3. Sample collection**

**WARNING!** Additional sample preparation before DNA extraction may be required!

#### **7.3.1 Nasopharyngeal, oropharyngeal smears**

Material is taken using special authorized medical devices according to the procedure established depending on the source of biological material.

Method limitations: local application of medicines (sprays, drops, creams and ointments) less than 24 hours before the test. When using aerosols and other forms of drugs for inhalation in the treatment of bronchial asthma, biomaterial should be taken no earlier than three hours after inhalation.

**WARNING!** Material is taken into tubes with **PREP-RAPID** reagent using a dry probe! It is necessary to exclude contact of the solution with skin, eyes and mucous membranes.

Take nasopharyngeal smear with a dry sterile probe, for which the probe is inserted with a slight movement along the outer wall of the nose to a depth of 2-3 cm to the lower nasal shell. Then slightly lower the probe to the bottom, enter the lower nasal passage under the lower nasal shell, make a rotary movement and remove along the outer wall of the nose.

Take oropharyngeal smear with a dry sterile probe, rotary movement from the surface of tonsils, palatine glands and the posterior pharyngeal wall.

#### **Order of sampling**

After taking the material transfer the probe into the tube with physiological saline solution, **PREP-RAPID** reagent, lysis solution or a transport medium authorized by the manufacturer, and rinse it thoroughly in the liquid for 10-15 s, avoiding splashing.

Remove the probe from the solution and press the probe against the inner wall of the tube above the solution level with a rotating motion to squeeze out excess liquid. Remove the probe completely from the tube and discard.

Close and tube tightly and mark it.

#### **7.3.2 Bronchoalveolar lavage; pleural fluid**

The material is taken in accordance with the instruction to the NA extraction reagent kit.

#### **In case of using PREP-NA and PREP-NA PLUS reagent kits:**

The material is taken into disposable, tightly screwed tubes with a volume of up to 50 mL. After taking the material, close the tube tightly and mark it.

Collect at least 500 µL of biomaterial into the tube. After taking the material, close the tube tightly and mark it.

### 7.3.3 Endotracheal, nasopharyngeal aspirate

The material is taken in accordance with the instruction to the NA extraction reagent kit.

#### In case of using PREP-NA and PREP-NA PLUS reagent kits:

The material is taken into disposable, tightly screwed tubes with a volume of up to 50 mL.

Collect at least 1.0 mL of biomaterial into the tube. After taking the material, close the tube tightly and mark it.

### 7.3.4 Phlegm

The material is taken in accordance with the instruction to the NA extraction reagent kit.

## 7.4. Transport and storage of the samples

Transport and storage conditions of biomaterial samples are stated in the instructions for use of the NA extraction reagent kits or the transport media used for transport and storage of samples.

It is allowed to transport and store biomaterial according to the conditions indicated in Table 5 (if it does not contradict the requirements stated in the instructions for use of the NA extraction reagent kits or the transport media used for transport and storage of samples):

Table 5. Biomaterial transport and storage conditions prior to DNA extraction

Biomaterial	Transport and storage temperature	Time period before DNA extraction
Bronchoalveolar lavage Endotracheal, nasopharyngeal aspirate Pleural fluid	From 2 °C to 8 °C	Up to 24 hours
	From minus 22 °C to minus 18 °C	Up to 7 days
Nasopharyngeal, oropharyngeal smears	From 2 °C to 8 °C	Up to 24 hours
	From minus 22 °C to minus 18 °C	Up to 1 month
Phlegm	From 18 °C to 25 °C	Up to 6 hours
	From 2 °C to 8 °C	Up to 3 days

**WARNING!** Only one freezing-thawing of the material is allowed.

## 7.5. Sample preparation (preparation of the suspension)

### 7.5.1. Nasopharyngeal, oropharyngeal smears; phlegm

Sample preparation (if necessary) is performed in accordance with the instructions for use for the NA extraction reagent kits.

### 7.5.2. Bronchoalveolar lavage; pleural fluid

7.5.2.1. Transfer 500 µL of biomaterial into a 1.5 mL single-use plastic tube.

7.5.2.2. Centrifuge the tube at RCF(g) 12,000 – 16,000 for 10 minutes.

7.5.2.3. Remove supernatant leaving approximately 50 µL (precipitate + liquid fraction) in the tube.

7.5.2.4. Add 500 µL of physiological saline solution to the precipitate.

7.5.2.5. Centrifuge the tube at RCF(g) 12,000 – 16,000 for 10 minutes.

7.5.2.6. Remove supernatant leaving approximately 100 µL (precipitate + liquid fraction) in the tube.

7.5.2.7. Close the tubes tightly.

Sample is ready for DNA extraction.

### 7.5.3. Nasopharyngeal, endotracheal aspirate

7.5.3.1. Transfer 1.0 mL of biomaterial into a 1.5 mL single-use plastic tube.

7.5.3.2. Centrifuge the tube at RCF(g) 12,000 – 16,000 for 10 minutes.

7.5.3.3. Remove supernatant leaving approximately 100 µL (precipitate + liquid fraction) in the tube.

7.5.3.4. Close the tubes tightly.

Sample is ready for DNA extraction.

## 8. PROCEDURE

### DNA extracting from biological material

DNA extraction is carried out according to the instruction to the NA extraction reagent kit.

Table 6. Kits recommended for DNA extraction (manufactured by “DNA-Technology”):

Reagent kit	Biomaterial	Minimal eluate volume, µL
PREP-NA	Nasopharyngeal, oropharyngeal smears; bronchoalveolar lavage; endotracheal, nasopharyngeal aspirate; phlegm; pleural fluid	50
	Nasopharyngeal, oropharyngeal smears (according to shortened method in Annex A)	
PREP-NA PLUS	Nasopharyngeal, oropharyngeal smears; bronchoalveolar lavage; endotracheal, nasopharyngeal aspirate; phlegm; pleural fluid	300
	Nasopharyngeal, oropharyngeal smears (according to shortened method in Annex A)	
PREP-RAPID	Nasopharyngeal, oropharyngeal smears	500
PREP-MB RAPID	Nasopharyngeal, oropharyngeal smears	100
PREP-OPTIMA	Nasopharyngeal, oropharyngeal smears	400
PREP-MB RAPID II	Nasopharyngeal, oropharyngeal smears	100

**WARNING!** Simultaneously with DNA extraction from biological material a negative control must be prepared and carried through all the stages of sample preparation. Physiological saline solution can be used as a negative control sample in volumes as indicated in the instructions for use of extraction kits or negative control sample that is include in the corresponding extraction kit.

#### 8.1 Preparing PCR for package S:

**WARNING!** The reagents and tubes should be kept away from direct sun light.

**WARNING!** When using package S, strips, strictly observe the completeness of the strips and caps. Do not use the caps for the strips of the other kits!

8.1.1 Mark one tube/stripped tube with the paraffin sealed PCR-mix for each test sample, negative control (C-), positive control (C+).

**WARNING!** The volume of reagents is calculated for no more than 12 runs assuming a variable number of test samples, 1 negative control and 1 positive control per run.

**Example:** To test 4 samples, mark 4 tubes for samples, one negative control tube “C-” and one positive control tube “C+”. Total number of tubes is 6.

8.1.2 Shake the tubes with Taq-polymerase solution on vortex for 3-5 seconds and spin on vortex for 1-3 seconds.

8.1.3 Add 10 µL of Taq-polymerase solution to each tube. Avoid paraffin layer break.

**WARNING!** When using the Rotor-Gene Q detecting thermal cycler for PCR, no mineral oil is added to the tubes!

8.1.4 Add one drop of mineral oil (~20 µL) to each tube. Cover the tubes/strips loosely with caps.

8.1.5 Shake the tube with positive control on vortex mixer and spin down the drops for 1-3 seconds.

**WARNING!**

1. Before introducing DNA preparation and negative control into tubes with PCR-mix, fulfill the recommendations for DNA preparation use listed in the NA extraction reagent kit instruction for use.
2. In case of using **PREP-RAPID, PREP-NA, PREP-NA PLUS** extraction kits, shake the tubes with DNA preparation and negative control for 3-5 seconds on vortex mixer and spin down the drops for 1-3 seconds.
3. In case of using **PREP-MB RAPID** DNA extraction kit, it is necessary to carefully, without shaking, spin the tubes with DNA preparation and negative control on vortex for 1-3 seconds and place the tubes into magnetic rack. If, after extraction, the supernatant containing the extracted DNA has been transferred to new tubes, shake the tubes with the DNA preparation and negative control “C-” on vortex for 3-5 seconds and spin down the drops for 1-3 seconds.
4. To prevent contamination, only open the caps of the tubes into which the sample is to be added and close them before adding the next sample. If strips are used, close the strip caps after adding the sample before proceeding with the next sample. Close the tubes/strips tightly. Use filter tips.

8.1.6 Add 5.0 µL of DNA preparation into corresponding tubes. Do not add DNA into the tubes “C-”, and “C+”. Avoid paraffin layer break.

8.1.7 Add 5.0 µL of negative control (C-) which passed whole DNA extraction procedure into “C-” tube. Avoid paraffin layer break.

8.1.8 Add 5.0 µL of positive control sample (C+) into corresponding tube. Avoid paraffin layer break.

8.1.9 Spin the tubes/strips for 3-5 seconds in a vortex mixer (if using Rotor-Gene Q thermal cycler, spinning is not required).

8.1.10 Set the tubes/strips into the real-time thermal cycler.

8.1.11 For DT instruments: Launch the operating software for DT instrument<sup>2</sup>. Add corresponding test<sup>3</sup>, specify the number and IDs of the samples, negative and positive controls. Specify position of the tubes/strips in thermal unit (see 8.1.10) and run PCR. See Table 7.

8.1.12 For Rotor-Gene Q, CFX96 and Applied Biosystems QuantStudio 5 thermal cyclers: perform PCR considering reaction mixture volume of 35 µL according to amplification programs shown in Table 8, 9,10.

---

<sup>2</sup> - Please, apply to Operation Manual for DTprime and DTlite Real-Time PCR instruments PART II.

<sup>3</sup> - Instructions for uploading "files with test parameters" can be found on "DNA-Technology's" website <https://www.dna-technology.com/assaylibrary>.

Table 7. The PCR program for DTlite, DTprime and DTprime II thermal cyclers for package S

Step	Temperature, °C	Min	Sec	Number of cycles	Optical measurement		Type of the step
1	80	0	30	1			Cycle
	94	1	30				
2	94	0	30	5			Cycle
	64	0	15		v		
3	94	0	10	45			Cycle
	64	0	15		v		
4	94	0	5	1			Cycle
5	25 <sup>1</sup>	...	...	Holding			Holding
v- optical measurements							
<sup>1</sup> – holding at 10°C is allowed							

Table 8. The PCR program for Rotor-Gene Q thermal cyclers (package S, tubes)

Cycling	Temperature	Hold Time	Cycle Repeats
Cycling 1	80 deg	60 sec	1 time
	94 deg	90 sec	
Cycling 2	94 deg	30 sec	5 times
	57 deg v	15 sec	
Cycling 3	94 deg	10 sec	45 times
	57 deg v	15 sec	
v– optical measurements, set Acquiring on Green (Fam), Yellow (Hex) detection channels at 57 °C			

Table 9. The PCR program for CFX96 thermal cyclers (packages S, U)

Step	Temperature, °C	Time, min:sec	Number of cycles (repeats)
1	80	01:00	1
2	94	01:30	1
3	94	00:15	50
4	64 √	00:20	
√ - optical measurements (Plate Read), set the fluorecence measurement on the Fam, Hex channels at 64 °C			

Table 10. The PCR program for Applied Biosystems QuantStudio 5 thermal cyclers (packages S, U)

Stage	Step	Temperature, °C	Time min: sec	Number of cycles (repeats)
Holding	1	80	01:00	1
	2	94	01:30	1
PCR	1	94	0:20	50
	2	64 v	0:20	
v- data collection for fluorophores (Fam, Vic (Hex)) is on				

## 8.2 Preparing PCR for package U, manual dosing

**WARNING!** The reagents and tubes should be kept away from direct sun light.

**WARNING!** For amplification use 0.2 mL single-use amplification tubes or 96-well PCR microplates<sup>4</sup>, sealed hermetically with thermal film. It is not recommended to use strips due to postamplification contamination hazard.

8.2.1 Mark the required number of 0.2 mL tubes or a 96-well PCR microplate for each test sample, negative control (C-) and positive control (C+).

Note: It is recommended to test at least 5 samples per test (3 test samples, negative control and positive control).

**Example:** to test 4 samples, mark 4 tubes/reserve 4 wells for samples, 1 tube/well for “C-” and 1 tube/well for “C+”. The resulting number of tubes/wells is 6.

8.2.2 Shake the tube with PCR-mix on vortex for 3-5 seconds, then spin in vortex for 1-3 seconds to collect the drops.

8.2.3 Add 6.0 µL of PCR-mix to each tube/well (including “C-” and “C+”).

8.2.4 Vortex the tube with PCR-buffer and TechnoTaq MAX polymerase for 3-5 seconds, then spin for 1-3 seconds to collect the drops.

**WARNING!** Take TechnoTaq MAX polymerase out from the freezer immediately prior to use.

8.2.5 Prepare the mixture of PCR-buffer and TechnoTaq MAX polymerase. Add into the one tube:

6.0 x (N+1) µL of PCR-buffer,

0.3 x (N+1) µL of TechnoTaq MAX polymerase,

where N is the quantity of samples to be tested taking to account “C-”, “C+”.

**Example:** to test 4 samples, “C-” and “C+” in one PCR run, mark 6 tubes/reserve 6 wells (4 tubes/wells for test samples, 1 tube/well for “C-” and 1 tube/well for “C+”). Prepare the mixture of PCR-buffer and Taq-polymerase for 7 (6+1) tubes/wells. Mix 42 µL of PCR-buffer and 2.1 µL of TechnoTaq MAX polymerase.

8.2.6 Shake the tube with the mixture of PCR-buffer and TechnoTaq MAX polymerase on vortex for 3-5 seconds, then spin on vortex for 1-3 seconds to collect the drops.

**WARNING!** Mixture of PCR-buffer and TechnoTaq MAX polymerase must be prepared immediately prior to use.

8.2.7 Add 6.0 µL of PCR-buffer and TechnoTaq MAX polymerase mixture into each tube/well with PCR-mix. Cover the tubes loosely.

**WARNING!** Follow the steps listed in pp. 8.2.8 – 8.2.14 within two hours after adding PCR-buffer and

<sup>4</sup> - 96-well plates are not used with DTLite detecting thermal cycler

TechnoTaq MAX polymerase mixture to PCR-mix.

8.2.8 Vortex the tube with positive control "C+" for 3-5 seconds and spin down the drops for 1-3 seconds.

**WARNING!**

1. Before introducing DNA preparation and negative control into tubes with PCR-mix, fulfill the recommendations for DNA preparation use listed in the NA extraction reagent kit instruction for use.
  2. In case of using **PREP-RAPID**, **PREP-NA**, **PREP-NA PLUS** extraction kits, shake the tubes with DNA preparation and negative control for 3-5 seconds on vortex mixer and spin down the drops for 1-3 seconds.
  3. In case of using **PREP-MB RAPID** DNA extraction kit, it is necessary to carefully, without shaking, spin the tubes with DNA preparation and negative control on vortex for 1-3 seconds and place the tubes into magnetic rack. If, after extraction, the supernatant containing the extracted DNA has been transferred to new tubes, shake the tubes with the DNA preparation and negative control "C-" on vortex for 3-5 seconds and spin down the drops for 1-3 seconds.
  4. To prevent contamination, only open the caps of the tubes into which the sample is to be added and close them before adding the next sample. Close the tubes tightly. Use filter tips.
- 8.2.9 Add 6.0 µL of DNA preparation into corresponding tubes/wells. Do not add DNA into the "C-", "C+" tubes/wells.
- 8.2.10 Add 6.0 µL of negative control (C-) which passed whole DNA extraction procedure into the corresponding tube/well.
- 8.2.11 Add 6.0 µL of positive control sample (C+) into the corresponding tube/well.
- 8.2.12 In case of using 96-well PCR microplates:**
- 8.2.12.1. Place the plate carefully, without shaking into the DTpack sealing device.
  - 8.2.12.2. Seal the PCR microplate with polymer thermal film according to the DTpack operation manual.
  - 8.2.12.3. Centrifuge the plate at RCF(g) 100 for 30 seconds.
- 8.2.13 In case of using tubes:**
- Spin the tubes for 3-5 seconds on vortex mixer (if using Rotor-Gene Q thermal cycler, spinning is not required).
- 8.2.14 Set the tubes/PCR microplate into the real-time thermal cycler.
- 8.2.15 Launch the operating software for DT instrument<sup>5</sup>. Add corresponding test<sup>6</sup>, specify the number and IDs of the samples, positive and negative controls. Specify position of the samples in thermal unit (see 8.2.14) and run PCR. See Table 11.
- 8.2.16 For CFX96, Applied Biosystems QuantStudio 5 and Rotor-Gene Q thermal cyclers perform PCR considering the volume of reaction mixture of 18 µL. See Table 9, 10, 12.

---

<sup>5</sup> - Please, apply to Operation Manual for DTprime and DTlite Real-Time PCR instruments PART II.

<sup>6</sup> - Instructions for uploading "files with test parameters" can be found on "DNA-Technology's" website <https://www.dna-technology.com/assaylibrary>.



Table 11. The PCR program for DTlite, DTprime and DTprime II thermal cyclers for package U

Step	Temperature, °C	Min	Sec	Number of cycles	Optical measurement	Type of the step
1	80	0	5	15		Cycle
	94	0	5			
2	94	5	00	1		Cycle
3	94	0	30	5		Cycle
	64	0	15		√	
4	94	0	10	45		Cycle
	64	0	15		√	
5	94	0	5	1		Cycle
6	25 <sup>1</sup>			Holding		Holding
√ - optical measurements						
<sup>1</sup> - holding at 10 °C is allowed						

Table 12. The PCR program for Rotor-Gene Q thermal cyclers (package U)

Cycling	Temperature	Hold Time	Cycle Repeats
Cycling 1	80 deg	60 sec	1 time
	94 deg	300 sec	
Cycling 2	94 deg	30 sec	5 times
	57 deg √	15 sec	
Cycling 3	94 deg	10 sec	45 times
	57 deg √	15 sec	
√– optical measurements, set Acquiring on Green (Fam), Yellow (Hex) detection channels at 57 °C			

### 8.3 Preparing PCR for package U, using DTStream (only for DTprime \*X\*, DTprime II \*X\*)

**WARNING!** The reagents and tubes should be kept away from direct sun light.

**WARNING!** For amplification use 384-well microplates hermetically sealed with thermal film.

Note – It is recommended to test at least 5 samples in 1 run (3 test samples, negative control and positive control).

8.3.1 Vortex the tube with PCR-mix for 3-5 seconds, then spin in vortex for 1-3 seconds to collect the drops.

8.3.2 Vortex the tube with PCR-buffer and TechnoTaq MAX polymerase for 3-5 seconds, then spin for 1-3 seconds to collect the drops.

**WARNING!** TechnoTaq MAX polymerase should be got out from the freezer immediately prior to use.

8.3.3 Prepare the mixture of PCR-buffer and TechnoTaq MAX polymerase according to the software for DTstream.

8.3.4 Vortex the tube with the mixture PCR-buffer and TechnoTaq MAX polymerase for 3-5 seconds, then spin for 1-3 seconds to collect the drops.

- 8.3.5 Vortex the tubes with positive control for 3-5 seconds and spin in vortex for 1-3 seconds to collect the drops.

**WARNING!**

1. Before dosing DNA preparation and negative control, fulfill the recommendations for DNA preparation use listed in the NA extraction reagent kit instruction for use.
  2. In case of using **PREP-RAPID**, **PREP-NA**, **PREP-NA PLUS** extraction kits, shake the tubes with DNA preparation and negative control for 3-5 seconds on vortex mixer and spin down the drops for 1-3 seconds.
  3. In case of using **PREP-MB RAPID** DNA extraction kit, it is necessary to carefully, without shaking, spin the tubes with DNA preparation and negative control on vortex for 1-3 seconds and place the tubes into magnetic rack. If, after extraction, the supernatant containing the extracted DNA has been transferred to new tubes, shake the tubes with the DNA preparation and negative control "C-" on vortex for 3-5 seconds and spin down the drops for 1-3 seconds.
- 8.3.6 Set the tubes with PCR-mix, the mixture of PCR-buffer and TechnoTaq MAX polymerase, DNA preparations, positive and negative controls and PCR microplate on the DTstream working table and conduct dosage of the components according to DTstream user manual.
- 8.3.7 After the end of dosing program on DTstream put the PCR microplate without shaking on the working table of DTpack sealing device.
- 8.3.8 Run the process of sealing of PCR microplate according to the user manual of DTpack sealing device.
- 8.3.9 Centrifuge the PCR microplate at RCF(g) 100 for 30 seconds.
- 8.3.10 Set the PCR microplate into the real-time thermal cycler.
- 8.3.11 Launch the operating software for DT instrument<sup>7</sup>. Add corresponding test<sup>8</sup>, specify the number and IDs of the samples, positive and negative controls. Specify position of the samples in thermal unit (see 8.3.10) and run PCR. See Table 11.

## 9. CONTROLS

The **Streptococcus pneumoniae REAL-TIME PCR Detection Kit** contains positive control C+.

Positive control is a cloned part of the bacterial genome. It is produced with genetic engineering techniques and characterized by automatic sequencing.

The PCR-mix from the kit includes the internal control (IC). IC is an artificial plasmid intended to assess the quality of PCR performance.

To reveal possible contamination a negative control is required.

**WARNING!** A negative control sample should go through all stages of DNA extraction. Physiological saline solution or negative control sample from an extraction kit can be used as a negative control sample in volumes indicated in supplied instructions.

For **Streptococcus pneumoniae REAL-TIME PCR Detection Kit** the test result is considered valid when:

- the exponential growth of the fluorescence level for the specific product is present, in this case the internal control is not considered;
- the exponential growth of the fluorescence level is absent for the specific product and is present for internal control.

---

<sup>7</sup> - Please, apply to Operation Manual for DTprime and DTlite Real-Time PCR instruments PART II.

<sup>8</sup> - Instructions for uploading "files with test parameters" can be found on "DNA-Technology's" website <https://www.dna-technology.com/assaylibrary>.

For ***Streptococcus pneumoniae* REAL-TIME PCR Detection Kit** the test result is considered invalid when the exponential growth of the fluorescence level for the specific product and for internal control is not observed.

If positive control (C+) does **not** express growing fluorescence of the specific product or positive result, it is required to repeat the whole test. It may be caused operation error or violation of storage and handling.

If negative control (C-) expresses growing fluorescence of the specific product or positive result, all tests of the current batch are considered false. Decontamination is required.

## 10. DATA ANALYSIS

**10.1** Registration of the results is carried out automatically during amplification by the software provided with detecting thermocycler.

**10.2** When using CFX96 detection thermal cyclers, use regression type analysis (Cq Determination Mode: Regression). In the "Baseline Subtraction" tab select "Baseline Subtraction Curve Fit".

**10.3** Result interpretation is carried out according to Table 13. The results are valid if the conditions for the interpretation of results obtained for control samples are met.

Table 13. PCR results interpretation

Detection channel		Result interpretation
Fam, (DNA), Cp/Cq/Ct	Hex, (IC), Cp/Cq/Ct	
Test samples		
Specified	Not considered	<i>Streptococcus pneumoniae</i> DNA is detected
Not specified	Specified	<i>Streptococcus pneumoniae</i> DNA is not detected
Not specified	Not specified	Unreliable result
Negative control		
Not specified	Specified	Negative result Run results are valid
Positive control		
Specified	Not considered	Positive result Run results are valid

**10.4** Unreliable result may be due to the presence of inhibitors in the NA preparation obtained from biological material; incorrect execution of the analysis protocol, noncompliance with the amplification temperature regime, etc. In this case it is necessary to repeat PCR with the available DNA preparation, or to re-extract DNA and perform PCR for this sample, or to re-collect biological material from the examinee (performed sequentially).

**10.5** If positive result is obtained for negative control, the results of the whole run are considered unreliable. In this case, special measures for identification and elimination of possible contamination are required.

**10.6** If negative result is obtained for positive control, the results of the whole run are considered unreliable. In this case, a repeated amplification of the whole batch of samples is required.

## 11. SPECIFICATIONS

### a. Analytical specificity

For biomaterial samples containing DNA of *Streptococcus pneumoniae*, thermal cycler software must register positive amplification result for specific products (*Streptococcus pneumoniae* genome fragments) on Fam/Green detection channel.

For biomaterial samples not containing DNA of *Streptococcus pneumoniae*, thermal cycler software must register negative amplification result for specific products (*Streptococcus pneumoniae* genome fragments) on Fam/Green detection channel and positive amplification result for internal control on Hex/Yellow/Vic detection channel.

There were no cross nonspecific reactions for any component included in the kit towards any other target of the test system.

There were no nonspecific positive amplification results in the presence of RNA of Influenza virus A(H3N2), Influenza virus A(H1N1pdm09), Influenza B virus, Human Coronavirus 229E, Human Coronavirus HKU-1, Human Coronavirus NL-63, Human Coronavirus OC-43, Human Metapneumovirus, Human Parainfluenza virus type 2, Human Parainfluenza virus type 3, Human Parainfluenza virus type 4, Human Parainfluenza virus type 1, Human Rhinovirus, MERS-CoV, Respiratory syncytial virus, SARS-CoV-2, DNA of Human Adenovirus, Human Bocavirus, *Bordetella parapertussis*, *Bordetella pertussis*, *Chlamydomonas pneumoniae*, *Haemophilus influenzae*, *Klebsiella pneumoniae*, *Moraxella catarrhalis*, *Mycoplasma pneumoniae*, *Staphylococcus aureus* (methicillin-resistant), *Legionella pneumophila*, as well as human DNA in concentration of  $1.0 \times 10^8$  copies per mL of sample.

### b. Analytical sensitivity (limit of detection)

Limit of detection is 5 copies of *Streptococcus pneumoniae* DNA per amplification tube.

Limit of detection was established via analysis of laboratory control's serial dilutions.

Limit of detection corresponds to the following DNA concentration values when using the specified NA extraction kits and the end elution (dilution) volumes of the extracted DNA:

Biomaterial	DNA extraction kit	Preparation volume, $\mu\text{L}$	Limit of detection, copies per sample
Nasopharyngeal, oropharyngeal smears in 500 $\mu\text{L}$ of transport medium <sup>9</sup>	<b>PREP-NA</b>	50	50
	<b>PREP-NA PLUS</b>	300	300
	<b>PREP-RAPID</b>	500	500
	<b>PREP-MB RAPID</b>	100	100
	<b>PREP-OPTIMA</b>	400	400
	<b>PREP-MB-RAPID II</b>	100	100
Bronchoalveolar lavage, phlegm, pleural fluid (extraction from 500 $\mu\text{L}$ of sample), endotracheal, nasopharyngeal aspirate (extraction from 1.0 mL of sample)	<b>PREP-NA</b>	50	50
	<b>PREP-NA PLUS</b>	300	300

<sup>9</sup> - **STOR-F** transport medium was used

### c. Diagnostic characteristics

Type of biomaterial	Diagnostic sensitivity	Diagnostic specificity
Nasopharyngeal smear	100% (95% CI: 86.28% – 100%)	100% (95% CI: 86.28% – 100%)
Oropharyngeal smear	100% (95% CI: 86.28% – 100%)	100% (95% CI: 86.28% – 100%)
Bronchoalveolar lavage	100% (95% CI: 86.28% – 100%)	100% (95% CI: 86.28% – 100%)
Endotracheal aspirate	100% (95% CI: 86.28% – 100%)	100% (95% CI: 86.28% – 100%)
Nasopharyngeal aspirate	100% (95% CI: 86.28% – 100%)	100% (95% CI: 86.28% – 100%)
Phlegm	100% (95% CI: 86.28% – 100%)	100% (95% CI: 86.28% – 100%)
Pleural fluid	100% (95% CI: 86.28% – 100%)	100% (95% CI: 86.28% – 100%)
	100% (95% CI: 97.91% – 100%)	100% (95% CI: 97.91% – 100%)

### d. Reproducibility and repeatability

Reproducibility is 100%.

Repeatability is 100%.

**NOTE** - The claimed specifications are guaranteed when DNA extraction is performed with **PREP-NA**, **PREP-NA PLUS**, **PREP-RAPID**, **PREP-MB RAPID**, **PREP-OPTIMA**, and **PREP-MB-RAPID II** extraction kits.

## 12. TROUBLESHOOTING

Table 14. Troubleshooting

	Result	Possible cause	Solution
C+	-	Operation error PCR inhibition Violation of storage and handling requirements	Repeat whole test Dispose current batch
C-	+	Contamination	Dispose current batch Perform decontamination procedures
IC	Invalid	PCR inhibition	Repeat whole test Resample

If you face to any undescribed issues contact our customer service department:

E-mail: [hotline@dna-technology.ru](mailto:hotline@dna-technology.ru)

<https://dna-technology.com/support>

### 13. QUALITY CONTROL

The quality control procedures performed in accordance with ISO 9001:2015 and ISO 13485:2016:

- observation of quality management in manufacturing of products;
- creation of values for customers;
- maintenance of the best service quality and customer management.

Contact our customer service with quality issues of **Streptococcus pneumoniae REAL-TIME PCR Detection Kit**.

#### **Technical support:**

E-mail: [hotline@dna-technology.ru](mailto:hotline@dna-technology.ru)

<https://www.dna-technology.com>

#### **Manufacturer:**

"DNA-Technology Research & Production", LLC,

142281, Russia, Moscow Region,

Protvino, Zheleznodorozhnaya Street, 20

Phone/fax: +7(495) 640.17.71

E-mail: [info@dna-technology.com](mailto:info@dna-technology.com)

<https://www.dna-technology.com>

**Seller:** "DNA-Technology" LLC,

117587, Russia, Moscow,

int. ter. Municipal District Chertanovo Severnoye,














Varshavskoye shosse, 125 Zh, building 5, floor 1, office 12;

Phone/fax: +7(495) 640.17.71

E-mail: [info@dna-technology.com](mailto:info@dna-technology.com)

<https://www.dna-technology.com>

#### 14. KEY TO SYMBOLS

	For research use only		Date of manufacture
	Temperature limit		Consult instructions for use
	Contains sufficient for<n>tests		Catalogue number
	Use-by date		Manufacturer
	Batch code		Keep away from sunlight
	Version		Non-sterile
	Caution		

**REF**

R1-P412-S3/4ER  
R1-P412-23/4ER  
R1-P412-UA/9ER

**VER**

1130.2025.06.23

## Annex A

### Shortened method of DNA extraction from biomaterial (nasopharyngeal, oropharyngeal smears) using PREP-NA, PREP-NA PLUS reagent kits

#### WARNING!

1. Before starting work:
  - preheat the thermostat to 65 °C;
  - take out the NA extraction reagent kit from the refrigerator and check that there is no precipitate in the lysis solution. In case of precipitation heat the vial with lysis solution on thermostat preheated to 65 °C to dissolve the precipitate completely. Then stir the solution by turning the vial upside down 5-10 times, avoiding foaming. Before use, cool the solution to room temperature (18 °C to 25 °C). The precipitate can also be dissolved at room temperature (18 °C to 25 °C) for approximately 12 hours.
2. Tube caps may open during heating! Use tubes with locking caps (e.g. Eppendorf Safe-Lock Tubes) or programmable thermostats with a clamp lid (e.g. solid-state programmable small-size thermostat TT-1-DNA-Technology, manufactured by "DNA-Technology R&P", LLC.).

#### Procedure:

1. Mark a 1.5 mL plastic tube for negative control (C-).
2. Enter 300 µL of lysis solution into each marked tube with 100 µL of biomaterial and into the "C-" tube. Do not touch the edges of the tube.
3. Add 100 µL of negative control into the "C-" tube.
4. Close the tubes tightly and shake on vortex for 3-5 seconds.
5. Heat the tubes on thermostat at 65 °C for 5 minutes.
6. Spin the tubes on vortex for 3-5 seconds.
7. Add 400 µL of precipitation buffer into each tube without touching the edges of the tube, close the tubes and shake on vortex for 3-5 seconds
8. Centrifuge the tubes at RCF(g) 12,000-16,000 at room temperature (from 18 °C to 25 °C) for 10 minutes.
9. Remove supernatant fully, using separate tip for each tube. Do not touch the precipitate.
10. Add 500 µL of wash solution No. 1 to the precipitate, close the tubes and mix the contents by carefully turning the tubes upside down 3-5 times.
11. Centrifuge the tubes at RCF(g) 12,000-16,000 at room temperature (from 18 °C to 25 °C) for 1 minute.
12. Remove supernatant fully, using separate tip for each tube. Do not touch the precipitate.
13. Add 300 µL of wash solution No. 2 to the precipitate, close the tubes and mix the contents by carefully turning the tubes upside down 3-5 times.
14. Centrifuge the tubes at RCF(g) 12,000-16,000 at room temperature (from 18 °C to 25 °C) for 1 minute.
15. Remove supernatant fully, using separate tip for each tube. Do not touch the precipitate. It is allowed to leave up to 20-30 µL of liquid above the precipitate.
16. Open the tubes and dry the precipitate at 65 °C for 5 minutes.
17. Add **50 µL (PREP-NA)** or **300 µL (PREP-NA PLUS)** of dilution buffer to the precipitate, shake the tubes on vortex for 3-5 seconds and spin down the drops for 3-5 seconds.



18. Heat the tubes on thermostat at 65 °C for 5 minutes. Shake the tubes on vortex for 3-5 seconds.
19. Centrifuge the tubes at RCF(g) 12,000-16,000 for 30 seconds at room temperature (from 18 °C to 25 °C) to spin down the condensate.

DNA preparation can be stored at temperature from minus 22 °C to minus 18 °C for no longer than one month or at from minus 72 °C to minus 68 °C for no longer than one year.

Before using the DNA preparation for PCR thaw the DNA preparation and negative control at room temperature (from 18 °C to 25 °C) or at temperature from 2 °C to 8 °C, then shake the tubes with DNA preparation and negative control on vortex for 3-5 seconds and spin on vortex for 1-3 seconds.

**WARNING!** It is only allowed to thaw DNA preparation once!

DNA preparation is ready for PCR.